

GX-AE SERIES

GX-A SERIES

GF-A SERIES

Multi-Function Balance

INSTRUCTION MANUAL

G X - A E series

GX-124AE/GX-224AE/GX-324AE

G X - A series

GX-124A/GX-224A/GX-324A

GX-203A/GX-303A/GX-403A/GX-603A/GX-1003A/GX-1603A

GX-2002A/GX-3002A/GX-4002A/GX-6002A/GX-10002A

GX-6001A/GX-10001A

G F - A series

GF-124A/GF-224A/GF-324A

GF-123A/GF-203A/GF-303A/GF-403A/GF-603A/GF-1003A/GF-1603A

GF-1202A/GF-2002A/GF-3002A/GF-4002A

GF-6002A/GF-10002A/GF-6001A/GF-10001A

The logo for A&D Company, Ltd. consists of the letters 'A' and 'D' in a bold, stylized, sans-serif font. The 'A' and 'D' are connected at the top, and the ampersand is integrated into the design between them.

A&D Company, Ltd.

1WMPD4003475B

© 2019 A&D Company Ltd. All rights reserved.

No part of this publication may be reproduced, transmitted, transcribed, or translated into any language in any form by any means without the written permission of A&D Company Ltd.

The contents of this manual and the specifications of the instrument covered by this manual are subject to change for improvement without notice.

Windows, Word and Excel are registered trademarks of the Microsoft Corporation.

Contents

| | |
|--|----|
| 1. Introduction | 5 |
| 1-1 Features | 5 |
| 1-2 About The Models..... | 6 |
| 1-3 About The GX-AE Series..... | 6 |
| 1-4 Compliance..... | 7 |
| 1-5 About Communication Manual | 7 |
| 2. Part Names, Installation And Precautions | 9 |
| 2-1 Installing The Balance | 13 |
| 2-2 Precautions..... | 13 |
| 2-3 During Use..... | 14 |
| 2-4 After Use | 15 |
| 2-5 Power Supply | 15 |
| 3. Display Symbols And Key Operation | 16 |
| 4. Weighing | 18 |
| 4-1 Units | 18 |
| 4-2 Basic Operation | 23 |
| 4-3 Counting Mode (PCS) | 25 |
| 4-4 Percent Mode (%)..... | 27 |
| 4-5 Animal Weighing Mode (Hold Function) | 28 |
| 5. Impact Detection Function | 28 |
| 5-1 Recording Impact History | 29 |
| 5-2 Output Impact History..... | 29 |
| 6. Response Adjustment / Self Check Function | 30 |
| 6-1 Response Adjustment | 31 |
| 6-2 Self Check Function / Automatic Setting Of Minimum Weighing Value By ECL | 31 |
| 6-2-1 For Balance Software Version 1.200 To 1.220..... | 32 |
| 6-2-2 For Balance Software Version 1.300 Or Later | 33 |
| 7. Calibration..... | 35 |
| 7-1 Automatic Calibration (GX-AE/GX-A Series Only) | 36 |
| 7-2 One-Touch Calibration (GX-AE/GX-A Series Only)..... | 37 |
| 7-3 Calibration Test Using An Internal Mass (GX-AE/GX-A Series 0.0001g Models Only)..... | 38 |
| 7-4 Calibration Using An External Weight | 39 |
| 7-5 Calibration Test Using An External Weight..... | 40 |
| 7-6 How To Set The External Weight Value..... | 41 |

| | |
|--|------------|
| 7-7 Correcting The Internal Mass Value Of The GX-AE/GX-A Series | 42 |
| 7-7-1 Correcting The Internal Mass Value Of The GX-AE/GX-A Series (Auto)..... | 43 |
| 7-7-2 Correcting The Internal Mass Value Of The GX-AE/GX-A Series (Manual) | 44 |
| 8. Function Switch And Initialization | 46 |
| 8-1 Permit Or Inhibit | 46 |
| 8-2 Initializing The Balance | 49 |
| 9. Function Table..... | 50 |
| 9-1 Setting The Function Table..... | 50 |
| 9-2 Details Of The Function Table | 52 |
| 9-3 Description Of The Class Environment Display | 59 |
| 9-4 Description Of The Data Output | 60 |
| 9-5 Description Of The Data Format | 60 |
| 9-6 Output Example Of The Data Format..... | 60 |
| 9-7 Clock And Calendar Function..... | 61 |
| 9-8 Comparator Function..... | 63 |
| 9-9 Description Of Application | 70 |
| 9-10 Output The Function Settings..... | 71 |
| 10. ID Number And GLP Report..... | 73 |
| 10-1 Main Objective..... | 73 |
| 10-2 Setting The ID Number..... | 73 |
| 10-3 GLP Report..... | 74 |
| 11. Data Memory..... | 79 |
| 11-1 Data Memory For Weighing Data | 79 |
| 11-2 Data Memory For Calibration And Calibration Test | 83 |
| 12. Statistical Calculation Mode | 86 |
| 12-1 How To Use The Statistical Calculation..... | 86 |
| 12-2 Statistical Calculation Mode (Example Of Use)..... | 91 |
| 13. Flow Measurement | 93 |
| 13-1 How To Use Flow Measurement | 93 |
| 14. Gross Net Tare Function | 97 |
| 14-1 Preparation Of Gross Net Tare Function | 97 |
| 14-2 Example Of Using The Gross Net Tare Function | 99 |
| 15. Minimum Weighing Warning Function..... | 100 |
| 15-1 Minimum Weighing Comparison | 101 |

| | |
|--|-----|
| 15-2 Input And Output Of The Minimum Weighing Value..... | 101 |
| 15-2-1 Setting From The Function Setting (The Balance Software Version 1.200 To 1.220)..... | 101 |
| 15-2-2 Setting From The Weighing Display (Balance Software Version 1.200 To 1.220)..... | 102 |
| 15-2-3 Setting From The Function Setting (Balance Software Version 1.300 Or Later) | 102 |
| 15-2-4 Setting From The Weighing Mode (Balance Software Version 1.300 Or Later) | 105 |
| 15-2-5 Batch Output The Minimum Weighing Value (Balance Software Version 1.300 Or Later) | 105 |
| 15-3 Setting Measurement Tolerance Of Minimum Weighing Value..... | 107 |
| 15-4 Data Output When Less Than Minimum Weighing Value | 108 |
| 16. Underhook | 109 |
| 17. Programmable-Unit..... | 110 |
| 18. Density (Specific Gravity) Measurement | 112 |
| 19. Password Lock Function | 117 |
| 19-1 Balance Software Version 1.200 | 117 |
| 19-2 Balance Software Version 1.211 Or Later | 118 |
| 19-3 Enable Password Lock Function | 119 |
| 19-4 How To Input The Password At The Start Of Weighing..... | 120 |
| 19-5 How To Logout..... | 121 |
| 19-6 Registering Password (Changing) | 122 |
| 19-7 Changing Password | 123 |
| 19-8 How To Delete The Password (USER 01 to 10)..... | 124 |
| 19-9 Missing Password | 124 |
| 20. Repeatability Check Function (GX-AE/GX-A Series Only) | 125 |
| 21. Interface Specification (Standard) | 126 |
| 22. Maintenance | 126 |
| 22-1 Treatment Of The Balance | 126 |
| 23. Troubleshooting | 126 |
| 23-1 Checking The Balance Performance And Environment..... | 126 |
| 23-2 Error Codes | 128 |
| 23-3 Other Display..... | 131 |
| 23-4 Asking For Repair..... | 131 |
| 24. Connection With Peripheral Device | 131 |
| 24-1 Command..... | 131 |
| 24-2 Key Lock Function..... | 131 |
| 25. How To Check The Software Version Of The Balance | 131 |
| 26. Specifications | 132 |

| | |
|--|------------|
| 26-1 Common Models | 132 |
| 26-1-1 Function | 132 |
| 26-1-2 Dimension | 132 |
| 26-2 Individual Models..... | 133 |
| 26-2-1 0.0001g Models..... | 133 |
| 26-2-2 0.001g Models..... | 134 |
| 26-2-3 0.01g Models..... | 135 |
| 26-2-4 0.1g Model..... | 136 |
| 27. External Dimensions | 137 |
| 27-1 Options And Peripheral Instruments..... | 140 |

1. Introduction

This manual describes how the GX-AE/GX-A/GF-A series balance works and how to get the most out of it in terms of performance. Read this manual thoroughly before using the balance and keep it at hand for future reference.

Behavior may differ depending on the software version of your balance.

For confirmation of the software version of the balance, refer to "25. How To Check The Software Version Of The Balance".

1-1 Features

- The balance has a self-check function that inspects the balance itself using electronically controlled load (ECL) and evaluates performance. Read this manual thoroughly before using the balance and keep it at hand for future reference.
- The balance can detect impact applied to its mass sensor and display the level of that impact. ISD (Impact Shock Detection).
- Continuous change of the balance can be calculated as flow rate, displayed and output. FRD : (Flow Rate Display).
- The balance is equipped with a data memory function, which can record weighing value, calibration result, and multiple unit mass (mass per sample in counting mode) (Up to 200 items are stored for weighing value).
- The GX-AE/GX-A series has automatic self calibration using the internal mass, adapting to temperature changes, setting time and interval time.
- Good Laboratory Practice (GLP) / Good Manufacturing Practice (GMP) data can be output using the RS-232C serial interface when making the sensitivity adjustment or sensitivity calibration adjustment. It is possible to record the results using an optional printer.
- A built-in clock and calendar that can add the time and date to the output data. (Setting and changing of the time can be limited to only an administrator by using the password lock function.)
- Comparator Indicators, displaying the comparison results with . (Depending on the setting, 5-step comparison is also possible.)
- Capacity Indicator, displaying the weight value in percentage relative to the weighing capacity.
- Hold Function, provided for weighing a moving object such as an animal.
- Underhook, for measuring density and weighing magnetic materials.
- Using the key lock function, key operation of the balance is disabled and operations can be made by commands from an external device only.
- Users of the balance can be limited by setting a password (Password lock function).
- The balance is equipped with an RS-232C serial interface and a USB interface to communicate with a computer. Windows computer using the Windows communication tools software (WinCT) make building a system very easy. The latest Win-CT software can be downloaded from the A&D website.
Windows is the registered trademark of the Microsoft Corporation.
- A small breeze break is included with the model featuring a readability of 0.001g.
- A glass breeze break is included with the model featuring a readability of 0.0001g.

1-2 About The Models

There are many models in the GX-AE/GX-A/GF-A series with differences in the models being their readability and weighing capacity. In this manual, they are listed collectively by the readability as shown in the table below.

| Model | Readability | Applicable model | | |
|---------------|-------------|----------------------------------|---|---|
| | | Internal mass with ionizer type | Internal mass type | General type |
| 0.0001g model | 0.0001g | GX-124AE GX-224AE GX-324AE | GX-124A GX-224A GX-324A | GF-124A GF-224A GF-324A |
| 0.001g model | 0.001g | | GX-203A / GX-303A / GX-403A / GX-603A / GX-1003A / GX-1603A | GF-123A/ GF-203A / GF-303A /GF-403A / GF-603A /GF-1003A / GF-1603A |
| 0.01g model | 0.01g | | GX-2002A / GX-3002A / GX-4002A / GX-6002A / GX-10002A | GF-1202A/ GF-2002A / GF-3002A /GF-4002A / GF-6002A /GF-10002A |
| 0.1g model | 0.1g | | GX-6001A / GX-10001A | GF-6001A / GF-10001A |

- For the GX-AE/GX-A series, a weight for sensitivity adjustment is built in. It is possible to use functions such as calibration and auto calibration using the internal mass.
- For the GF-A series, sensitivity adjustment weights are not built-in. When calibrating, it is necessary to prepare an external weight.

1-3 About The GX-AE Series

- An ionizer (static eliminator), which causes no breeze, is built into the breeze break. The ionizer can eliminate static electricity from the weighing sample before weighing, reducing error due to static electricity. (Static is eliminated by bipolar ions generated by direct-current corona discharge being projected onto the sample.)
- The ionizer electrodes are designed to be removed for cleaning and replacement.
- An IR (touchless) switch is attached to the ionizer, and static elimination can be started without touching ionizer.
- Power is supplied from the balance to allow the ionizer to be operated without using an AC adapter.
- PRINT or RE-ZERO or the static elimination function of the ionizer can be operated by using the optional foot switch (AX-SW137-PRINT or AX-SW137-REZERO).
- A removable glass breeze break is equipped as standard.
- As an option board is installed, so other options (GXA-03/04/06/09/17/23/24/25/26, FXi-08, etc.) cannot be used.
- For instructions on using the ionizer and IR switch, download the instruction manual for "GXA-17 Large Glass Breeze Break with Ionizer" from our website (<https://www.aandd.jp>) and refer to it.

Static electricity

In general, when the ambient humidity is less than 45%RH, nonconductors such as powders, paper, and plastics easily become charged with static electricity. The influence of static electricity may cause a weighing error of several milligrams. The ionizer neutralizes this electrical charge.

1-4 Compliance

Compliance with FCC Rules

Please note that this equipment generates, uses and can radiate radio frequency energy. This equipment has been tested and has been found to comply with the limits of a Class A computing device pursuant to Subpart J of Part 15 of FCC rules. These rules are designed to provide reasonable protection against interference when equipment is operated in a commercial environment. If this unit is operated in a residential area, it may cause some interference and under these circumstances the user would be required to take, at his own expense, whatever measures are necessary to eliminate the interference.

(FCC = Federal Communications Commission in the U.S.A.)

Compliance With Directives of CE mark

CE This device features radio interference suppression, safety regulation and restriction of Hazardous Substances in compliance with the following Council Directives

Council directive 2014/30/EU EN61326 EMC directive

Council directive 2014/35/EU EN60950 Safety of Information Technology Equipment

Council directive 2011/65/EU EN50581 Restriction of the use of certain Hazardous Substances

The CE mark is an official mandatory European marking.

Please note that any electronic product must comply with local laws and regulations when sold or used anywhere outside Europe.

1-5 About Communication Manual

About the contents of the communication, download "Communication manual" from our website (<https://www.aandd.jp>) and refer to it.



A & D Instruments Ltd. hereby declare that the following Weighing product conforms to the requirements of the council directives on ...

**Electromagnetic Compatibility (EMC) 2014/30/EU,
Low Voltage Equipment (LVD) 2014/35/EU and
Restriction of the use of certain Hazardous Substances (RoHS) 2011/65/EU**

provided that they bear the CE mark of conformity.

Model/Series....GX-A/GF-A Series

Standards applicable:

EN 61326-1:2013

Electrical equipment for measurement, control and laboratory use -EMC requirements Part 1: General requirements

EN 60950-1:2006+A11:2009+A1:2010+A12:2011+A2:2013

Safety of Information Technology Equipment

EN 50581:2012

Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances

CE Mark first applied 16 October 2017

Signed for A&D Instruments in Oxford England 27 November 2017


J. Ghuman
Managing Director


2. Part Names, Installation And Precautions

The balance is a precision instrument. Unpack it carefully. The packing contents depend on the balance model. See the illustrations to confirm that everything is included. When options are combined at time of shipping, optional accessories may be included.


Keep the packing material to be used for transporting the balance in the future.

GX-AE / GX-A / GF-A 0.0001g models

GX-A/GF-A
Glass Breeze Break



GX-AE
Glass Breeze Break with Ionizer



Pull out the left and right locking handles.

Breeze break ring

Weighing pan


Bubble spirit level

Display


Keys

Main unit


Leveling foot



Main unit cover
(PET resin)




USB cable (approx. 1.8m)



Press the left and right locking handles to secure the breeze break to the balance.


Rear of main unit

Each model that includes a glass breeze break with ionizers (GX-124AE / GX-224AE / GX-324AE) has an ionizer connector, IR sensor connector and external key jack.




Connecting the glass breeze break with ionizer and the IR sensor

Type: GX-124AE / GX-224AE / GX-324AE



- Insert the cable extending from the rear of the breeze break into the ionizer connector at the rear of the balance.
- Insert the IR sensor plug into the IR sensor connector at the rear of the balance.




Attach the AC adapter label to the adapter as shown in the illustration.

Note

- Please confirm that the AC adapter type is correct for your local voltage and receptacle type.
- Please use the specified dedicated AC adapter for the balance.
- Do not use the included AC adapter for models that are not considered compatible with the AC adapters.
- If you use the wrong AC adapter, the balance and other equipment may not operate properly.


GX-A / GF-A 0.001g models



Note

- Please confirm that the AC adapter type is correct for your local voltage and receptacle type.
- Please use the specified dedicated AC adapter for the balance.
- Do not use the included AC adapter for models that are not considered compatible with the AC adapters.
- If you use the wrong AC adapter, the balance and other equipment may not operate properly.

GX-A / GF-A 0.01g/0.1g models



Note

- Please confirm that the AC adapter type is correct for your local voltage and receptacle type.
- Please use the specified dedicated AC adapter for the balance.
- Do not use the included AC adapter for models that are not considered compatible with the AC adapters.
- If you use the wrong AC adapter, the balance and other equipment may not operate properly.

2-1 Installing The Balance

Install the balance as follows:

1. Refer to "2-2 Precautions" for installing the balance.
2. Assemble the balance as shown in the illustration above.
3. Adjust the leveling feet to level the balance. Confirm it using the bubble spirit level.
4. Confirm that the adapter type is correct for the local voltage and power receptacle type.
5. Connect the AC adapter to the balance. Warm up the balance for at least 30 minutes with nothing on the weighing pan.


2-2 Precautions

To get the optimum performance from the balance and acquire accurate weighing data, note the following:



- Install the balance in an environment where the temperature and humidity are not excessive. The best operating temperature is about $20^{\circ}\text{C}\pm 2^{\circ}\text{C}$ at about 45~60%RH relative humidity.
- Install the balance where it is free of dust.
- The weighing table should be solid, free from vibration and drafts, and as level as possible. (An anti-vibration table or stone table is ideal)
- Install the balance in a stable place avoiding vibration and shock. Corners of rooms on the first floor are best, as they are less prone to vibration.
- Install the balance where it is not affected by heaters or air conditioners.
- Install the balance where it is not exposed to direct sunlight.
- Install the balance away from equipment which produces magnetic fields.
- Level the balance by adjusting the leveling feet and confirm it using the bubble spirit level.
- Warm up the balance for at least 30 minutes. Plug in the AC adapter as usual.
- Calibrate the balance before use or after having moved it to another location. Refer to "7. Calibration".

Caution

Do not install the balance where flammable or corrosive gas is present.





How to adjust the bubble spirit level



2-3 During Use




To acquire accurate weighing data, note the following:

- Discharge static electricity from the material to be weighed. When a sample could have a static charge, the weighing data is influenced. If the ambient humidity becomes 45% or less, insulators such as plastics are liable to become static electricity. Ground the balance and try the following.
 - An anti-static treatment has been applied to the breeze break components.
 - Eliminate the static electricity by GXA-25, AD-1683 as an accessory.
 - Or try to keep highly the ambient humidity
 - Or use a metal shield case.
 - Or wipe a charged plastic sample with the wet cloth.
- The breeze break (0.001g models only) and the clear main unit cover are provided as accessories. The breeze break components may be charged with static electricity when they are unpacked or when the humidity is low. If the weighing value is unstable or the balance has a problem with repeatability, remove the breeze break. Or wipe the clear plates with a moistened cloth, use an accessory DC static eliminator, GXA-25, AD-1683 or apply an anti-static spray.
- This balance uses a strong magnet as part of the balance assembly, so please use caution when weighing magnetic materials such as iron. If there is a problem, use the underhook



on the bottom of the balance to suspend the material away from the influence of the magnet.

- Cancel the temperature difference between a sample and the environment. When a sample is warmer (cooler) than the ambient temperature, the sample will be lighter (heavier) than the true weight. This error is due to a rising (falling) draft around the sample.
- Make each weighing gently and quickly to avoid errors due to changes in the environmental conditions.
- When placing a sample on a weighing pan, do not give a strong shock or do not exceed the weighing capacity. And place in the center.
- Do not drop things upon the weighing pan, or place a sample on the pan that is beyond the balance weighing capacity. Place a sample in the center of the weighing pan.
- Do not use a sharp instrument such as a pencil to press the keys. Use your finger only.
- Press the **RE-ZERO** key before each weighing to prevent possible errors.
- Take into consideration the affect of air buoyancy on a sample when more accuracy is required.
- Keep the balance interior free of dust and foreign materials.



2-4 After Use

- Avoid mechanical shock to the balance.
- Do not disassemble the balance. Contact the local A&D dealer if the balance needs service or repair.
- Do not use organic solvents to clean the balance. Clean the balance with a lint free cloth that is moistened with warm water and a mild detergent.
- Avoid dust and water so that the balance weighs correctly. Protect the internal parts from liquid spills and excessive dust.


2-5 Power Supply

- When the AC adapter is connected, the balance is in the standby mode if the standby indicator is on. This is a normal state and does not harm the balance. For accurate weighing, keep the AC adapter connected to the balance and AC power unless the balance is not to be used for a long period of time. Please warm up the balance for at least 30 minutes (one hour for 0.0001g models).


3. Display Symbols And Key Operation

Display symbols

- Number of statistical data (Statistical calculation mode)
- Displays the weight data relative to the weighing capacity, in percentage, in the weighing mode (**Capacity indicator**)




Blinking display contents




Key operation

Key operations affect how the balance functions. Normal key operation during measurement is “Press and release the key immediately” or “Press the key”. Please do not “Press and hold the key (for 2 seconds)” unless is required.









Press the key

(Press and release the key immediately.)



Press and hold the key

for 2 seconds.


| Key | When pressed and released | When pressed and held (for 2 seconds) |
|---|---|--|
|  | Turns the display ON:OFF . The standby indicator is displayed when the display is turned off. The weighing mode is enabled when the display is turned on. When password function is enable, password input display will be displayed. Refer to "19-4 How to Input The Password At The Start Of Weighing" This ON:OFF key is available anytime. Pressing the ON:OFF key during operation will interrupt operation and turn the display OFF. * | |
|  | In the weighing mode, turns the digit for readability on and off. In the counting or percent mode, enters the sample storing mode. | <ul style="list-style-type: none"> • Enters the function table mode. Please refer to "9. Function Table". • Runs the repeatability check function when pressed and held for another 2 seconds after the function table menu is displayed. Please refer to "20. Repeatability Check Function". (GX-AE / GX-A series only) |
|  | Switches the weighing units stored in the function table. Refer to "4. Weighing". Displaying of the unit mg is available for 0.0001g models only. | Displays other items of the calibration menu. Please refer to "6-2 Self Check Function/ Automatic Setting Of Minimum Weighing Value by ECL". |
|  | Performs calibration of the balance using the internal mass. (GX-AE/GX-A series) | Displays other items of the calibration menu. |
|  | Stores the weighing data in memory or outputs to a printer or personal computer depending on the function table settings. (Factory setting = output) | Enters mode to change the unit mass registration number in counting mode. By changing the function table: <ul style="list-style-type: none"> • Outputs "Title block" and "End block" for GLP, GMP report. • Displays the data memory menu. • Enters mode for reading density number in flow measurement. |
|  | Sets the display to zero. | |

* When the "Gross net tare function" is selected, the display is turned off by pressing and holding (for 2 seconds). Please refer to "14. Gross Net Tare Function".

4. Weighing

4-1 Units

With the GX-AE/GX-A/GF-A series balance, the following weighing units and weighing modes are available:
 Note: The unit "mg" is available for the 0.0001g models only. "mg" is displayed after "g" on 0.0001g models.



Counting mode

Percent mode

Density mode (To use this mode, it must be stored in the function table as described on page 21. For details about this mode, refer to "18. Density (Specific Gravity) Measurement". To select this mode, press the **MODE** key until the processing indicator blinks with the unit "g" displayed. "DS" appears only when the density value is displayed.)

Programmable-unit (No unit displayed. For details, refer to "17. Programmable-Unit".)

A unit or mode can be selected and stored in the function table as described on page 21.
 If a weighing mode (or unit of weight) has been turned off, that mode or unit will be missing in the sequence. Tael has four varieties, one of which can be selected and installed at the factory.

To select a unit or mode for weighing, press the **MODE** key.

For details about the units and modes, see the table below:

| Name (unit, mode) | Abbrev. | Display | Function table (Storing mode) | Conversion factor 1 g = |
|----------------------------------|---------|-------------------------------------|-------------------------------|-----------------------------------|
| Gram | g | g | g | 1 g |
| Milligram | mg | mg | mg | 0.001 g |
| Counting mode | PCS | PCS | PCS | — |
| Percent mode | % | % | % | — |
| Ounce (Avoir) | OZ | OZ | OZ | 28.349523125 g |
| Pound | Lb | Lb | Lb | 453.59237 g |
| Pound/Ounce | L OZ | L OZ | LO | 1Lb=16 oz, 1 oz=28.349523125 g |
| Troy Ounce | OZt | OZt | OZt | 31.1034768 g |
| Metric Carat | ct | ct | ct | 0.2 g |
| Momme | mom | mom | mom | 3.75 g |
| Pennyweight | dwt | dwt | dwt | 1.55517384 g |
| Grain (UK) | GN | GN | GN | 0.06479891 g |
| Tael (HK general, Singapore) | TL | TL | TL | 37.7994 g |
| Tael (HK jewelry) | | | | 37.429 g |
| Tael (Taiwan) | | | | 37.5 g |
| Tael (China) | | | | 31.25 g |
| Tola (India) | tol | tol | tol | 11.6638038 g |
| Messghal | MES | MES | MES | 4.6875 g |
| Density mode (See note below) | DS | DS is used to show the density. | DS | — |
| Programmable-unit (Multi-unit) | MLT | MLT | MLT | — |

Note: The blinking processing indicator with “g” indicates that the density mode is selected.

The tables below indicate the weighing capacity and the readability for each unit, depending on the balance model.

| Unit | GX-124AE | GX-224AE | GX-324AE | Readability |
|------------------------------|----------|----------|----------|-------------|
| | GX-124A | GX-224A | GX-324A | |
| | GF-124A | GF-224A | GF-324A | |
| | Capacity | | | |
| Gram | 122 | 220 | 320 | 0.0001 |
| Milligram | 122000 | 220000 | 320000 | 0.1 |
| Ounce (Avoir) | 4.30 | 7.05 | 10.50 | 0.00001 |
| Troy Ounce | 3.92 | 6.43 | 9.64 | 0.00001 |
| Metric Carat | 610 | 1000 | 1500 | 0.001 |
| Momme | 32.5 | 53.3 | 80.0 | 0.0001 |
| Pennyweight | 78.4 | 128 | 192 | 0.0001 |
| Grain (UK) | 1882 | 3086 | 4629 | 0.002 |
| Tael (HK general, Singapore) | 3.22 | 5.29 | 7.93 | 0.00001 |
| Tael (HK jewelry) | 3.25 | 5.34 | 8.01 | 0.00001 |
| Tael (Taiwan) | 3.25 | 5.33 | 8.00 | 0.00001 |
| Tael (China) | 3.90 | 6.40 | 9.60 | 0.00001 |
| Tola (India) | 10.4 | 17.1 | 25.7 | 0.00001 |
| Messghal | 26.0 | 42.6 | 64.0 | 0.0001 |

| Unit | | GX-203A | GX-303A | GX-403A | GX-603A | GX-1003A | GX-1603A | Readability |
|------------------------------|------------|------------|-------------|-------------|------------|------------|------------|-------------|
| | GF-123A | GF-203A | GF-303A | GF-403A | GF-603A | GF-1003A | GF-1603A | |
| | Capacity | | | | | | | |
| Gram | 122 | 220 | 320 | 420 | 620 | 1100 | 1620 | 0.001 |
| Ounce (Avoir) | 4.30 | 7.76 | 11.28 | 14.81 | 21.86 | 38.80 | 57.14 | 0.00005 |
| Pound | 0.268 | 0.485 | 0.705 | 0.925 | 1.366 | 2.425 | 3.571 | 0.000005 |
| Pound/Ounce | 0Lb 4.30oz | 0Lb 7.76oz | 0Lb 11.28oz | 0Lb 14.81oz | 1Lb 5.86oz | 2Lb 6.80oz | 3Lb 9.14oz | 0.01oz |
| Troy Ounce | 3.92 | 7.07 | 10.28 | 13.50 | 19.93 | 35.36 | 52.08 | 0.00005 |
| Metric Carat | 610 | 1100 | 1600 | 2100 | 3100 | 5500 | 8100 | 0.005 |
| Momme | 32.5 | 58.6 | 85.3 | 112.0 | 165.3 | 293.3 | 432.0 | 0.0005 |
| Pennyweight | 78.4 | 141 | 205 | 270 | 398 | 707 | 1041 | 0.001 |
| Grain (UK) | 1882 | 3395 | 4938 | 6481 | 9568 | 16975 | 25000 | 0.02 |
| Tael (HK general, Singapore) | 3.22 | 5.82 | 8.46 | 11.11 | 16.40 | 29.10 | 42.85 | 0.00005 |
| Tael (HK jewelry) | 3.25 | 5.87 | 8.54 | 11.22 | 16.56 | 29.38 | 43.28 | 0.00005 |
| Tael (Taiwan) | 3.25 | 5.86 | 8.53 | 11.20 | 16.53 | 29.33 | 43.20 | 0.00005 |
| Tael (China) | 3.90 | 7.04 | 10.24 | 13.44 | 19.84 | 35.20 | 51.84 | 0.00005 |
| Tola (India) | 10.4 | 18.8 | 27.4 | 36.0 | 53.1 | 94.3 | 138.8 | 0.0001 |
| Messghal | 26.0 | 46.9 | 68.2 | 89.6 | 132.2 | 234.6 | 345.6 | 0.0005 |

| Unit | | GX-2002A | GX-3002A | GX-4002A | GX-6002A | GX-10002A | Readability |
|------------------------------|-------------|-------------|------------|------------|--------------|-------------|-------------|
| | GF-1202A | GF-2002A | GF-3002A | GF-4002A | GF-6002A | GF-10002A | |
| | Capacity | | | | | | |
| Gram | 1220 | 2200 | 3200 | 4200 | 6200 | 10200 | 0.01 |
| Ounce (Avoir) | 43.0 | 77.6 | 112.8 | 148.1 | 218.6 | 359.7 | 0.0005 |
| Pound | 2.68 | 4.85 | 7.05 | 9.25 | 13.66 | 22.48 | 0.00005 |
| Pound/Ounce | 2Lb 11.03oz | 4Lb 13.60oz | 7Lb 0.87oz | 9Lb 4.15oz | 13Lb 10.69oz | 22Lb 7.79oz | 0.01oz |
| Troy Ounce | 39.2 | 70.7 | 102.8 | 135.0 | 199.3 | 327.9 | 0.0005 |
| Metric Carat | 6100 | 11000 | 16000 | 21000 | 31000 | 51000 | 0.05 |
| Momme | 325 | 586 | 853 | 1120 | 1653 | 2720 | 0.005 |
| Pennyweight | 784 | 1414 | 2057 | 2700 | 3986 | 6558 | 0.01 |
| Grain (UK) | 18827 | 33951 | 49383 | 64815 | 95680 | 157410 | 0.2 |
| Tael (HK general, Singapore) | 32.2 | 58.2 | 84.6 | 111.1 | 164.0 | 269.8 | 0.0005 |
| Tael (HK jewelry) | 32.5 | 58.7 | 85.4 | 112.2 | 165.6 | 272.5 | 0.0005 |
| Tael (Taiwan) | 32.5 | 58.6 | 85.3 | 112.0 | 165.3 | 272.0 | 0.0005 |
| Tael (China) | 39.0 | 70.4 | 102.4 | 134.4 | 198.4 | 326.4 | 0.0005 |
| Tola (India) | 104 | 188 | 274 | 360 | 531 | 874 | 0.001 |
| Messghal | 260 | 469 | 682 | 896 | 1322 | 2176 | 0.005 |

| Unit | GX-6001A | GX-10001A | Readability |
|------------------------------|--------------|-------------|-------------|
| | GF-6001A | GF-10001A | |
| | Capacity | | |
| Gram | 6200 | 10200 | 0.1 |
| Ounce (Avoir) | 218 | 359 | 0.005 |
| Pound | 13.6 | 22.4 | 0.0005 |
| Pound/Ounce | 13Lb 10.69oz | 22Lb 7.79oz | 0.01oz |
| Troy Ounce | 199 | 327 | 0.005 |
| Metric Carat | 31000 | 51000 | 0.5 |
| Momme | 1653 | 2720 | 0.05 |
| Pennyweight | 3986 | 6558 | 0.1 |
| Grain (UK) | 95680 | 157410 | 2 |
| Tael (HK general, Singapore) | 164.0 | 269.0 | 0.005 |
| Tael (HK jewelry) | 165.0 | 272.0 | 0.005 |
| Tael (Taiwan) | 165.0 | 272.0 | 0.005 |
| Tael (China) | 198.0 | 326.0 | 0.005 |
| Tola (India) | 531.0 | 874.0 | 0.01 |
| Messghal | 1322 | 2176 | 0.05 |


Storing Units

The units or modes can be selected and stored in the function table. The sequence of displaying the units or modes can be arranged to fit the frequency of use.

The units stored are maintained in non-volatile memory, even if the AC adapter is removed.

Select a unit or mode and arrange the sequence of display as follows:


1. Press and hold the **SAMPLE** key until **bRSFnC** of the function table is displayed, then release the key.
2. Press the **SAMPLE** key several times to display **Unit**.
3. Press the **PRINT** key to enter the unit selection mode.
4. Specify a unit or mode in the order to be displayed using the following keys.
 - SAMPLE** key..... To sequentially display the units.
 - RE-ZERO** key To specify a unit or mode. The stabilization indicator **o** appears when the displayed unit or mode is specified. If the key is pressed in units already selected, the stability mark disappears.
5. Press the **PRINT** key to store the units or modes. The balance displays **End** and then displays the next menu of the function table.
6. Press the **CAL** key to exit the function table. Then the balance returns to the weighing mode with the selected unit.
7. To select other unit or mode for weighing, press the **MODE** key.



Unit setting example

The example below sets the units in the order with g (gram) as the first unit followed by pcs (counting mode).

1. Press and hold the **SAMPLE** key until **bR5FnC** of the function table is displayed, then release the key.
2. Press the **SAMPLE** key several times to display **Unit**.
3. Press the **PRINT** key to enter the unit selection mode.
4. Press the **RE-ZERO** key to specify the unit of g
The stabilization indicator **○** appears when the unit is specified.
5. Press the **SAMPLE** key to display **Unit PCS**.
6. Press the **RE-ZERO** key to specify the unit of pcs.
The stabilization indicator **○** appears when the unit is specified.
7. Press the **PRINT** key to store the units.
The balance displays **End** and then displays the next menu item of the function table.
8. Press the **CAL** key to exit the function table. Then the balance returns to the weighing mode with g, the unit selected first.
9. Press the **MODE** key to switch between g and pcs (g→pcs).




4-2 Basic Operation

1. Press **MODE** key, and then select the appropriate units
In this case, select "g".
2. Place a container on the weighing pan, if necessary.
Press the **RE-ZERO** key to cancel the weight (tare).
The balance displays **0.00 g**. (The decimal point position depends on the balance model.)
3. Place a sample on the pan or in the container. Wait for the stabilization indicator **o** to be displayed. Read the value.
4. Remove the sample and container from the pan.

Note

- Press the **SAMPLE** key to turn on or off the digit for the readability.
- The weighing data can be stored in memory. For details, refer to "11. Data Memory".
- When the **ON:OFF** key is pressed with a container placed on the weighing pan and weighing is started, the balance automatically cancels the weight (tare) and displays **0.00 g**.



About the operation at when power is turned on

The balance will determine the reference zero point when the **ON:OFF** key is pressed to enter the weighing mode.

Depending on the load condition at that time, it will automatically judge whether to perform zero or tare operation. The condition for determining which is used is "power on zero range", and when power on zero range is exceeded, the tare subtraction operation is performed.

About re-zero operation

By pressing the **RE-ZERO** key, the display can be changed to zero.

Re-zero with the **RE-ZERO** key will automatically determine whether zero or tare operation is performed.

The condition for determining which is used is "zero range", and when zero range is exceeded, the tare subtraction operation is performed.

About measurement range

For the balance, the range that can be weighed is determined by model.

The total amount (net amount + tare quantity) up to the maximum display of each model is displayed, and when the maximum display is exceeded, **E** is displayed to indicate that the weighing range is exceeded.

When in excess in negative, **-E** is displayed.

| Model | | Power on zero range | Zero range | -E display range |
|----------------------|-----------|---------------------|----------------|-----------------------|
| GX-124AE, GX-124A | GF-124A | Approx. ± 50g | Approx. ± 2g | Approx. -50g or less |
| GX-224AE, GX-224A | GF-224A | | Approx. ± 4g | |
| GX-324AE, GX-324A | GF-324A | | Approx. ± 6g | |
| | GF-123A | Approx. ± 100g | Approx. ± 2g | Approx. -100g or less |
| GX-203A | GF-203A | | Approx. ± 4g | |
| GX-303A | GF-303A | | Approx. ± 6g | |
| GX-403A | GF-403A | | Approx. ± 8g | |
| GX-603A | GF-603A | | Approx. ± 12g | |
| GX-1003A | GF-1003A | | Approx. ± 20g | |
| GX-1603A | GF-1603A | | Approx. ± 32g | |
| | GF-1202A | Approx. ± 1kg | Approx. ± 20g | Approx. -1kg or less |
| GX-2002A | GF-2002A | | Approx. ± 40g | |
| GX-3002A | GF-3002A | | Approx. ± 60g | |
| GX-4002A | GF-4002A | | Approx. ± 80g | |
| GX-6002A | GF-6002A | | Approx. ± 120g | |
| GX-10002A | GF-10002A | | Approx. ± 200g | |
| GX-6001A | GF-6001A | Approx. ± 1kg | Approx. ± 120g | Approx. -1kg or less |
| GX-10001A | GF-10001A | | Approx. ± 200g | |

4-3 Counting Mode (PCS)

This is the mode to determine the number of objects in a sample based on the standard sample unit mass. Unit mass means the mass of one sample. The smaller the variables in each sample unit mass is, the more accurate the counting will be. This series balance is equipped with the Automatic Counting Accuracy Improvement (ACAI) function to improve the counting accuracy.

Note

- * For counting, use samples that have a unit mass at least ten times greater than that of the readability in grams.
- * If the sample unit mass variable is too large, it may cause a counting error.
- * To improve the counting performance, use the ACAI function frequently or divide the samples into several groups and count each group.

Selecting the counting mode

1. Press the **MODE** key to select **PCS** (PCS = unit)

Storing a sample unit mass

2. Press the **SAMPLE** key to enter the sample unit mass storing mode. Even in the storing mode, pressing the **MODE** key will switch to the next mode.
3. To select the number of samples, press the **SAMPLE** key several times. It may be set to 5, 10, 25, 50 or 100.

Note

A greater number of samples will yield more accurate counting result.


4. Place a container on the weighing pan, if necessary.
Press the **RE-ZERO** key to cancel the weight (tare). The number specified in step 3 appears.
25.0 is displayed if 25 is selected in "3".
5. Place the number of samples specified on the pan. In this example, 25 pieces.
6. When **PRINT** key pressed, unit mass is stored and changes the count display. (Ex: when the number is 25, **25 PCS** is displayed.

Note

- * If the balance judges that the mass of the samples is too light to acquire accurate weighing, it displays an error requiring the addition of more samples to the specified number and press the **PRINT** key. When the unit mass is stored correctly, the balance proceeds to the counting mode.
- * If the balance judges that the mass of the samples is too light and is not adequate to be used as the unit mass, it displays **Lo**.
- * Registered unit mass is remembered even when the power is turned off.

Number mode (counting)

7. Counting is possible.




Counting Mode Using The ACAI Function

The ACAI is a function that improves the accuracy of the unit mass automatically by increasing the number of samples as the counting process.

ACAI: Automatic Counting Accuracy Improvement

After registering unit mass of "6", proceed to the following "8".

Processing mark




8. If a few more samples are added, the processing indicator turns on. To prevent an error, add three or more. The processing indicator does not turn on if overloaded. Try to add the same number of samples as displayed.
9. The balance re-calculates the unit mass while the processing indicator is blinking. Do not touch the balance or samples on the pan until the processing indicator turns off.
10. Counting accuracy is improved when the processing indicator turns off. Each time the above operation is performed, a more accurate unit mass will be obtained. There is no definite upper limit of ACAI range for the number of samples exceeding 100. Try to add the same number of samples as displayed.
11. Remove all the samples used in ACAI and proceed with the counting operation using the improved unit mass.

Note ACAI will not function on the unit mass entered using the keys, or digital input mode.

Storing the unit mass

By using the data memory function, 50 instances of storing a sample unit mass can be stored.

1. Set the function setting item "Data memory function (*DATA*)" to "Stores unit mass in counting (*DATA I*)". Refer to "9. Function Table".
2. The displayed " *P *** " is the selected unit mass registration number.
3. Press and hold the **PRINT** key for 2 seconds to switch to the mode to change the unit mass registration number.
 - RE-ZERO** key Changes the registration number(+)
 - MODE** key Changes the registration number (-)
 - PRINT** key Decides on the displayed registration number.
 - CAL** key Cancel the displayed registration number.
4. Multiple unit masses can be stored by registering them with different unit mass registration numbers.
 - P ***: The unit weight registration number is entered.



Note

- Unit weight can be read by " UN:mm " command.
(mm corresponds to P01 to P50 with 01 to 50.)
- The read unit mass can output by " ?UW " command and can be changed by " UW " command.

Caution

- * ACAI cannot be used for the read unit mass.

4-4 Percent Mode (%)

The percent mode displays the weighting value in percentage compared to a 100% reference mass and is used for target weighing or checking the sample variance.

Selecting The Percent Mode

1. Press the **MODE** key to select the unit **%** (Percent mode).

Storing The 100% Reference Mass

2. Press the **SAMPLE** key to enter the 100% reference mass storing mode.
Even in the storing mode, pressing the **MODE** key will switch to the next mode.
3. Place a container on the weighing pan, if necessary. Press the **RE-ZERO** key to cancel the weight (tare). The balance displays **100.00 %**.
4. Place the sample to be set as the 100% reference mass on the pan or in the container.
5. Press the **PRINT** key to store the reference mass.
The balance displays **100.00 %**. (The decimal point position depends on the reference value. The reference mass stored, even if the AC adapter is removed, is maintained in non-volatile memory.)

Note


- If the balance judges that the mass of the sample is too light to be used as a reference, it displays **Lo**.
- The displayed percentage is based on the 100% reference mass.

| Model | 100% mass | Decimal point position |
|---------------------|-------------------|------------------------|
| Readability 0.0001g | 0.0100g ~ 0.0999g | 1% |
| | 0.1000g ~ 0.9999g | 0.1% |
| | 1.0000g ~ | 0.01% |
| Readability 0.001g | 0.100g ~ 0.999g | 1% |
| | 1.000g ~ 9.999g | 0.1% |
| | 10.000g ~ | 0.01% |
| Readability 0.01g | 1.00g ~ 9.99g | 1% |
| | 10.00g ~ 99.99g | 0.1% |
| | 100.00g ~ | 0.01% |
| Readability 0.1g | 1.0g ~ 9.9g | 1% |
| | 10.0g ~ 99.9g | 0.1% |
| | 100.0g ~ | 0.01% |

- Registered values are stored even when the power is turned off.

Reading the percentage

6. Please a sample to be compared to the reference mass on the pan. The displayed percentage is based on the 100% reference mass.



4-5 Animal Weighing Mode (Hold Function)

This is the mode to weigh a moving object such as an animal, even when the display of the weighing data fluctuates. The hold function allows the average weight of the animal to be displayed. To use the hold function, set the function in the function table. Refer to "9. Function Table" and "9-3 Description Of The Class Environment Display" for details.

5. Impact Detection Function





The GX-AE/GX-A/GF-A series has a function to detect impact to the mass sensor section and to display the impact level.

By lowering the impact level at the time of loading, it is possible not only to alleviate variation in the weighing value but also to reduce the risk of failure of the mass sensor section.

Especially when incorporating the balance in a production line, etc. and weighing by means such as an automated system, impact to the sensor may be applied greater than expected.

When designing automatic systems and the like, it is recommended that you minimize the impact level as much as possible while checking the shock indicator.

Impact level display is from level 0 to level 4, 5 level.

| Impact level | Shock indicator | Buzzer | Contents |
|--------------|--|-----------|---------------------------------------|
| 0 | No | No | Safe |
| 1 | SHOCK  | No | Caution |
| 2 | SHOCK  | No | Caution : Consider impact mitigation |
| 3 | SHOCK  | One beep | Warning : Do not apply greater impact |
| 4 | SHOCK  | Two beeps | Danger : Sensor may be damaged |

With balance software version 1.300 or later, you can turn off the impact detection function by setting $bRSFnC/15d$ of the function settings to \square .

Even if the impact detection function is turned off, it records in the balance when there is a shock.

Note

- Impact on the weighing sensor may be applied to the weighing pan at time of loading, or it may be applied from the table on which the balance is installed.

The impact detection function also works for impact applied from the table.

5-1 Recording Impact History

Impacts of impact level 3 or higher are stored on the balance with data and time included (maximum 50 data instances).

When the password lock function is on (*Lock 1* or *Lock 2*), the login user information is added when outputting the impact history. (Balance software version 1.211 or later.)

Note

- If 50 data instances is exceeded, the data with the lowest impact level is overwritten.
- The stored impact history cannot be deleted.
- Impact data where the balance is not energized (during transport, etc.) is not stored.

5-2 Output Impact History

The stored impact history can be output by sending a specified command to the balance or performing a key operation.

Note


- The impact history format differs according to the software version of the balance.

Output by command

The stored impact data will be output all at once by sending a ?SA command to the balance.

Output by key operation

1. Press the **ON:OFF** key to turn off the display.
2. With the display off, press the **ON:OFF** key while holding down the **MODE** key.
3. **-L 15t-** is displayed, and the stored impact data is output all at once.



Impact history output example

The impact history format differs according to the software version of balance.

- For balance software version 1.200

Date, time, and impact level are each output on a separate line.

Output example

2018/05/29,11:08:18,SHOCK LV4

- For balance software version 1.211 or later

Date, time, impact level, login and login user information are output together on one line.

The login user information varies by the setting of the login user and the setting of Function table *Lock* when receiving impact.

| Output | Login user | Function table <i>Lock</i> |
|--------------|---------------|----------------------------|
| , --, | No login user | 0, 1, 2 |
| ,00, ADMIN | Administrator | 1 |
| ,01~10, USER | User | 1 |
| , --, GUEST | Guest | 2 |

Output example

2018/05/29,11:08:18,SHOCK LV,3, --,



2018/05/29,11:12:27,SHOCK LV,4,00,ADMIN


2018/05/29,11:13:38,SHOCK LV,3,01,USER

2018/05/29,11:17:04,SHOCK LV,4, -- ,GUEST

6. Response Adjustment / Self Check Function

This function stabilizes the weight value, reducing the influence on weighing that is caused by drafts and/or vibration at the place where the balance is installed. This function adjusts by automatically analyzing the environment or by hand-operation. The function has three stages as follows : Changing the weighing speed changes the display refresh rate.


| Display | Function setting | Response characteristic |
|---------|------------------|---|
| FAST | [ond 0] | Fast response, Sensitive value |
| MID. | [ond 1] |   |
| SLOW | [ond 2] | Slow response, Stable value |



6-1 Response Adjustment

Response adjustment can be changed by the following method.

1. Press and hold the **MODE** key for 2 seconds until **RESPONSE** is displayed, and then press the **MODE** key again.
 2. Press the **MODE** key to select a weighing speed. Either **FAST**, **MID** or **SLOW** can be selected.
 3. After a few seconds of inactivity the balance displays **End**. Then, it returns to the weighing mode and displays the updated response indicator.
- The response indicator remains displayed for a while.



Note

When setting the Response adjustment, "Condition (*Cond*)" and "Display refresh rate (*SPd*)" in the Function Table "Environment display (*bRSFnC*)" are changed as below.

| Display | <i>Cond</i> (Condition) | <i>SPd</i> (Display refresh rate) | <i>S-bb</i> (Stability band width) |
|---------|-------------------------|-----------------------------------|------------------------------------|
| FAST | 0 | 2 | 2 |
| MID. | 1 | 0 | 1 |
| SLOW | 2 | 0 | 1 |

When using a combination other than the above, please set individually as shown in "9. Function Table".

Note

If **RESPONSE** is displayed and you leave without pressing the **MODE** key, the "Self-check function" is activated. Please refer to "6-2 Self Check Function". For the setting method, refer to "9. Function Table".

6-2 Self Check Function / Automatic Setting Of Minimum Weighing Value By ECL

With the self check function confirmation and display of repeatability can be performed in addition to failure diagnosis, and whether or not the balance's performance is being exhibited can be easily checked. It is also possible to display and register the reference value of the minimum weighing value using repeatability data. For details of the minimum weighing value, refer to the technical information on our website. (<https://www.aandd.jp>)

6-2-1 For Balance Software Version 1.200 To 1.220

1. Press and hold the **[MODE]** key for 2 seconds while weighing is displayed.
2. Release the key when **[RESPONSE]** display blinks.
3. **[CH:.....]** displays and the self-diagnosis function is started.


"ECL" will be displayed in a few seconds. When the **[MODE]** key is pressed with **[CH:.....]** displayed, the change in weighing value for repeatability due to the electronically controlled load (ECL) is seen. (Applied from balance software version 1.100 or later)

4. When the diagnosis is completed, the diagnosis result is displayed. When there are no problems in the balance, the **[CH PASS]** display blinks. If **[CH FAIL]** is displayed blinking, there is a possibility that a fatal fault has occurred in the balance. Please request repair.

[SAMPLE] key ... It is possible to switch the display of diagnostic result, repeatability, minimum weighing value.

[PRINT] key ... The displayed contents are output. When repeatability is displayed, **OK** is lit up in the catalog specifications. When the catalog specification is exceeded, **Err** blinks and calls for review of the installation environment. (Applied from balance software version 1.100 or later)

5. Press the **[CAL]** key to display **[End]** and move to the display for minimum weighing value registration.
6. When registering as minimum weighing value of minimum weight value display warning function performs the following. Press the **[RE-ZERO]** key to change No/Go and display **[Go]**. Press the **[PRINT]** key. If not registered, press the **[PRINT]** key while displaying **[No]**, or press **[CAL]**. Return to the weighing display.




* For minimum weighing warning function, refer to " 15. Minimum Weighing Warning Function".

Note

For the self-check-function, the settings of the balance software versions 1.200 can only be made when logged in as administrator (AdMin) if the password lock function is turned on.

Setting procedure figure



Example of batch output

```

-MINIMUM WEIGHT-
      A & D
MODEL  GX-10002A
S/N    T2000112
ID     LAB-0123
DATE   2019/01/22
TIME   12:51:55
ECL
RESULT
1      +20.07 g
2      +20.06 g
3      +20.06 g
4      +20.06 g
5      +20.05 g
6      +20.06 g
7      +20.05 g
8      +20.06 g
9      +20.06 g
10     +20.06 g
SD     5.7 mg
TOLERANCE
0.10 %
MINIMUM WEIGHT
11.40 g
REMARKS
SIGNATURE
    
```

7. Calibration

Since the balance's resolution is high, weighing values may change due to gravity and daily environmental changes. It is necessary to perform calibration (sensitivity adjustment) with the weight in order to keep the weighing values from changing even if gravity or the environment changes.

It is recommended that you calibrate if the balance is installed for the first time or relocated, or when the weighing values change significantly in daily inspection, etc.

Adjustment means to adjust the weighing value of the balance using the reference weight or internal mass. Calibration is to weigh with the reference weight and compare how much the result deviates from the reference value. (Adjustment is not performed in calibration.)

Calibration (Sensitivity adjustment)

- Auto calibration Automatically adjust the balance using the internal mass depending on the temperature change of the operating environment or the set time and interval time. (GX-AE / GX-A series)
- Calibration using the internal mass Using the internal mass, adjust the balance with a single touch. (GX-AE / GX-A series)
- Calibration using an external weight Using an external mass, adjust the balance with an external mass.

Calibration test (Sensitivity calibration)

- Calibration test with an external weight Output the result of checking the accuracy of weighing using your own weight.
 - * No adjustment is made.
- Calibration test with an internal mass Output the result of checking the accuracy of weighing using the internal mass.
 - * No adjustment is made.(GX-AE / GX-A series with 0.0001g type)

Caution

- Do not allow vibration or drafts to affect the balance during calibration.
- To output the data for GLP/GMP using the RS-232C interface, set "GLP/GMP output (*INF0*)" of "Data output (*dOut*)". Refer to "9. Function Table". The time and date can be added to the GLP/GMP report. If the time or date is not correct, adjust them. Refer to "9-7 Clock and Calendar Function".
- By setting "Data memory (*dAtA*)" of the function table, the data of calibration (sensitivity adjustment execution record) and calibration test (sensitivity calibration) can be stored in memory.

Caution when using your external weight

- The accuracy of the weight used in calibration affects the accuracy of the balance after calibration.
- Select the mass to be used for calibration and calibration tests from the table below.

| Model | Usable calibration weight | Factory setting | Adjustable range |
|-------------------------------|---------------------------------------|-----------------|------------------------|
| GX-124AE, GX-124A, GF-124A | 50g, 100g | 100g | -0.9999g ~ +0.9999g |
| GX-224AE, GX-224A, GF-224A | 50g, 100g, 200g | 200g | |
| GX-324AE, GX-324A, GF-324A | 50g, 100g, 200g, 300g | | |
| GF-123A | 50g, 100g | 100g | -9.999g ~ +9.999g |
| GX-203A, GF-203A | 50g, 100g, 200g | 200g | |
| GX-303A, GF-303A | 50g, 100g ~ 300g (100g interval) | 200g | |
| GX-403A, GF-403A | 50g, 100g ~ 400g (100g interval) | 400g | |
| GX-603A, GF-603A | 50g, 100g ~ 600g (100g interval) | 500g | |
| GX-1003A, GF-1003A | 50g, 100g ~ 1000g (100g interval) | 1000g | |
| GX-1603A, GF-1603A | 50g, 100g ~ 1600g (100g interval) | 1000g | |
| GF-1202A | 500g, 1000g | 1000g | |
| GX-2002A, GF-2002A | 500g, 1000g, 2000g | 2000g | -99.99g ~ +99.99g |
| GX-3002A, GF-3002A | 500g, 1000g ~ 3000g (1000g interval) | | |
| GX-4002A, GF-4002A | 500g, 1000g ~ 4000g (1000g interval) | | |
| GX-6002A, GF-6002A | 500g, 1000g ~ 6000g (1000g interval) | | |
| GX-10002A, GF-10002A | 500g, 1000g ~ 10000g (1000g interval) | | |
| GX-6001A, GF-6001A | 500g, 1000g ~ 6000g (1000g interval) | | |
| GX-10001A, GF-10001A | 500g, 1000g ~ 10000g (1000g interval) | 10000g | +99.9g |

Display



This indicator means calibration data (sensitivity adjustment and sensitivity calibration) is being imported.

Do not allow vibration or drafts to affect the balance while the indicator is displayed.

7-1 Automatic Calibration (GX-AE/GX-A Series Only)

This function automatically calibrates the balance according to ambient temperature change, setting time and interval time. If GLP output is selected in the function table, the balance outputs the calibration report after the calibration.

- In the auto calibration mode, either the temperature change ([Fnc 0]), the setting time ([Fnc 1]), or the interval time ([Fnc 2]) can be set with the function setting [Fnc].
- For the setting time, the three function setting of [tME 1], [tME 2] and [tME 3] can be set.
- Interval time can be set from 0.5h to 24h with function setting [int].

Caution

If something is on the weighing pan, the balance judges that it is in use and does not perform automatic self calibration.


The criteria that the balance judges in use are as follows.

| | | | |
|-----------------|---------------|----------------|----------------|
| 0.0001g models | 0.001g models | 0.01g models | 0.1g models |
| Lower than 0.5g | Lower than 2g | Lower than 20g | Lower than 20g |

To maintain the accurate calibrated state, keep the weighing pan clear while not in use.



This mark blinking (◀) indicates that the automatic self calibration will start. If the balance is not in use, after blinking for a while, the balance will start automatic self calibration using the internal mass. The blinking duration depends on the environment.



This indicates that the balance is importing calibration data. Do not allow vibration or drafts to affect the balance while this indicator is displayed. After calibration, the balance returns to indicate the previous display.

Note The balance can be used while the indicator blinks. But, it is recommended that to maintain the accuracy, stop using the balance and confirm that there is nothing on the pan and allow the balance to perform self calibration.

Depending on the setting of "8. Function Switch And Initialization", "Change prohibited" or "Changeable (usable)" can be selected.

7-2 One-Touch Calibration (GX-AE/GX-A Series Only)

This function calibrates the balance using the internal mass.

1. Connect the AC adapter and warm up the balance for at least 30 minutes with nothing on the weighing pan.
2. Press the **CAL** key. The balance displays **CALⁱⁿ**.
3. The balance performs calibration using the internal mass. Do not allow vibration or drafts to affect the balance.
4. After the calibration, If the "GLP output (*inF₀*)" parameter of the function table is set, the balance outputs a sensitivity adjustment execution record.
5. The balance will automatically return to weighing mode after calibration.

About the internal mass

The value of the internal mass may change due to factors such as the operating environment and aging. Correct the internal mass value as necessary. Refer to "7-7 Correcting The Internal Mass Value Of The GX-AE/GX-A Series", "7-7-1 Correcting The Internal Mass Value Of The GX-AE/GX-A Series (Auto)", "7-7-2 Correcting The Internal Mass Value Of The GX-AE/GX-A Series (Manual).

Since the internal mass is about 200g, the possibility of error may increase as the weighing value increases.

To maintain the weighing accuracy, perform the calibration using an external weight periodically, as described below "7-4 Calibration Using An External Weight".

7-3 Calibration Test Using An Internal Mass (GX-AE/GX-A Series 0.0001g Models Only)


This function tests the weighing accuracy using an internal mass. (The result is output, but the sensitivity adjustment is not performed.)

Only high-precision analysis balances (0.0001g) can perform calibration test.

1. Connect the AC adapter and warm up the balance for at least one hour with nothing on the weighing pan.
2. Press and hold the **[CAL]** key for 2 seconds until **[CAL in]** is displayed.
3. The zero point is checked. Do not apply vibration.
4. The zero point checked is displayed.
5. Full scaling is checked. Do not apply vibration.
6. The full scaling checked is displayed. The reference values of each balance type are following. When the full scale display is within the measurement tolerance, it means that the sensitivity adjustment was performed correctly with the internal mass.

| Type | Reference value of full scaling | Measurement tolerance |
|----------|---------------------------------|-----------------------|
| GX-124AE | 100.0000g | ±0.2mg |
| GX-124A | | |
| GX-224AE | 200.0000g | |
| GX-224A | | |
| GX-324AE | | |
| GX-324A | | |


7. When the output is set to GLP/GMP report (**Info 1** or **2**), the calibration test result is output.
8. The valance automatically returns to weighing mode.



7-4 Calibration Using An External Weight

This function calibrates the balance using an external weight.


1. Connect the AC adapter and warm up the balance for at least 30 minutes with nothing on the weighing pan.
2. Press and hold the **CAL** key for 2 seconds until **CAL out** is displayed, then release the key.
3. Make sure that nothing is on the weighing pan and press the **PRINT** key to weigh the zero point.
Do not apply vibration, etc.
4. Place the external weight on the weighing pan and press the **PRINT** key. Do not apply vibration etc.
5. Remove the external weight from the weighing pan.
6. After calibration, if GLP output is set, "sensitivity adjustment execution record" is output or stored in data memory.
7. The display automatically returns to weighing display.
8. Place the external weight again and check that the set value is ± 2 digits.
If it is out of range, pay attention to the surrounding environment and start from "1".



7-5 Calibration Test Using An External Weight

This function tests the weighing accuracy using an external weight and outputs the result. This is available only when the GLP/GMP output parameter is set to (*dout, dAtA 2*). (Calibration test does not perform calibration)

1. Connect the AC adapter and warm up the balance for at least 30 minutes with nothing on the weighing pan.
2. Press and hold the **CAL** key for 2 seconds until **[[out** is displayed and release the key.
3. Make sure that nothing is on the weighing pan and press the **PRINT** key and weigh the zero point.
Do not apply vibration etc.
4. The weighing value of zero point is displayed for several seconds. Place the external weight on the weighing pan and press the **PRINT** key. Weigh the external weight.
Do not apply vibration, etc.
5. Weighing value of the external weight is displayed for several seconds.
6. Remove the external weight from the weighing pan.
7. The sensitivity calibration status is output or stored in the data memory.
8. It automatically returns to the weighing display.



7-6 How To Set The External Weight Value

When calibrating the balance or performing a calibration test, the external weight you have on hand can be set. (Refer to "Usable calibration weight" on Page 36.)

After **[CAL 0]** is displayed the external weight value can be set as shown in "7-4 Calibration Using An External weight". Or, after **[CC 0]** is displayed, the external weight value can be set as shown in "7-5 Calibration Test Using An External Weight".

1. From calibration **[CAL 0]** or calibration test **[CC 0]** display, press the **[SAMPLE]** key.
2. Using the **[RE-ZERO]** key, select the calibration weight (refer to page 36) while all digits are blinking.
3. Specify the calibration weight value by the following keys.


[SAMPLE] key…… Switches the display condition to
 "All digits blinking"
 (calibration weight selection mode) or
 "The last four digits blinking"
 (value adjustment mode).

[RE-ZERO] key … (+) In the adjustment range setting, the value becomes +9999 digits after -9999 digits.

[MODE] key …… (-) In the adjustment range setting, the value becomes -9999 digits after +9999 digits.

[PRINT] key …… Registers the changed external weight value. Registered values are stored even when the power is turned off.

[CAL] key Suspend setting. (Returns to **[CAL 0]** or **[CC 0]**.)



Ex:
 Updated the external weight
 3000.12g

7-7 Correcting The Internal Mass Value Of The GX-AE/GX-A Series

Internal mass value can be corrected with function setting [5].

There are two correction methods, as follows.

Auto…… This is a method of correcting the internal mass weight value based on an external weight.

Manual…… This is a method of correcting by digitally inputting a correction reference value (internal weight conversion value).

Note

- Correction of internal mass value can not be executed at factory setting.


Refer to "8. Function Switch And Initialization" or the following setting method, and enable changing of the function setting and correction the internal mass value.

Setting procedure


- Press the **ON:OFF** key to turn off the display.
- Hold down the **PRINT** and **SAMPLE** keys, and press the **ON:OFF** key to display **P5**.
- Press the **PRINT** key and set the "internal mass correction switch" and "function setting switch" to "1" with the next key.

SAMPLE key…… Select the switch (blinking digit).

RE-ZERO key…… Change the value of the blinking switch.



- Press the **PRINT** key to register and display the weighing display.




7-7-1 Correcting The Internal Mass Value Of The GX-AE/GX-A Series (Auto)

Calibrate referring to "7-4 Calibration Using An External Weight".

This is method of correcting the internal mass weight value based on an external weight.

After calibration with the external mass, the balance automatically loads and unloads the internal mass and corrects the internal mass value.


The corrected mass value is maintained in non-volatile memory even if the AC adapter is removed.



Setting procedure

The internal mass value cannot be corrected at factory settings. Refer to the Setting procedure in "7-7 Correcting The Internal Mass Value Of The GX-AE/GX-A Series" and enable changing of the function setting and correction the internal mass value.

1. In weighing mode, press and hold the **[SAMPLE]** key for 2 seconds to display **[bASFnC]**.
2. Press the **[SAMPLE]** key several times until **[CS in]** appears.
3. Press the **[PRINT]** key to display **[Auto]**.
4. When preparation is completed, press the **[PRINT]** key. **[CAL SET]** is displayed and the internal mass value is automatically corrected.
5. When adjustment of the internal mass value is completed, **[CAL in]** is displayed and calibration is performed automatically with the adjusted internal weight.
6. When calibration is completed, **[MANUAL]** is displayed.
7. Press the **[CAL]** key twice to return to weighing mode.
8. Make sure that the balance was corrected using the external weight for correcting. If it is not corrected properly, return to "1".
(Be aware of vibration during correcting)




7-7-2 Correcting The Internal Mass Value Of The GX-AE/GX-A Series (Manual)

The balance can correct the internal mass value within the range shown below. This function corrects the internal mass value to conform to an external weight. The corrected mass value is maintained in non-volatile memory even if the AC adapter is removed. The internal mass value is corrected as follows:

| Model | Target | Range | Model | Target | Range |
|----------|-----------|----------------------|----------|--------------------|---------------------|
| GX-124AE | 100.000g | $\pm 0.9999\text{g}$ | GX-203A | 200.000g | $\pm 9.999\text{g}$ |
| GX-124A | | | GX-303A | | |
| GX-224AE | GX-403A | | | | |
| GX-224A | GX-603A | | 500.000g | | |
| GX-324AE | 200.0000g | | GX-1003A | 1000.000g | |
| GX-324A | | | GX-1603A | | |
| | | | GX-2002A | 2000.00g | |
| | GX-3002A | | | | |
| | GX-4002A | | | | |
| | GX-6002A | | 5000.00g | | |
| | GX-10002A | | | | |
| | GX-6001A | | 5000.0g | $\pm 99.9\text{g}$ | |
| | GX-10001A | | | | |

Example: GX-6002A




After performing one touch calibration, place the external weight on the balance and check the correction amount.

(In the example, since there is deviation of -0.06g at 2000g, the correction amount for the GX-6002A is +0.15g/5kg because the correction target is 5000g)

Setting procedure

The internal mass can not be set by factory setting. Refer to "7-7 Correcting The Internal Mass Value Of The GX-AE/GX-A Series" to make it possible to change the internal setting and correct the internal mass value.

1. Press and hold the **SAMPLE** key for 2 seconds to display **bASFnC**.
(Enter the function setting)
2. Press the **SAMPLE** key several times until **[5 in** appears.
3. Press the **PRINT** key to display **Auto**.
4. Press the **SAMPLE** key to display **MANUAL** and press the **PRINT** key.
Select the following keys.
 - RE-ZERO** key(+) Select the correction value.
(After +9999 digits will be -9999 digits.)
 - MODE** key(-) Select the correction value.
(After -9999 digits will be +9999 digits.)
 - PRINT** key Register and display the following items.
 - CAL** key Cancel and display the following items.
5. Press the **CAL** key twice, to return to the weighing display.
6. Press the **CAL** key and calibrate with the internal mass.
7. Place the external weight on the balance check whether the value was corrected properly. If it is not corrected properly, return to "1". (Be aware of vibration during correcting)



8. Function Switch And Initialization

8-1 Permit Or Inhibit

The balance stores parameters that must not be changed unintentionally adjustment data for accurately weighing, data for adapting to the usage environment, data to control the communications interface, etc. "A function selection switch" is provided to protect those parameters and it can be used to select "Change prohibited" or "Changeable (usable)". By setting to "change prohibited", that function cannot be entered, so inadvertent change.

"Switch for function selection" has the following five.

"Function table", "Calibration using the internal mass", "Calibration using the external weight", "Automatic self calibration", "Internal mass correction".

1. Press the **ON:OFF** key to turn off the display.
2. While pressing and holding the **PRINT** key and the **SAMPLE** key, press the **ON:OFF** key to display **P5**.
3. Press the **PRINT** key. Then the balance displays the function switches.
 - SAMPLE** key To select a switch to change the parameter. The selected switch blinks.
 - RE-ZERO** key To change the parameter of the switch selected.
 - Change prohibited
 - |** Changeable (usable)
 - PRINT** key To store the new parameter and return to the weighing mode.
 - CAL** key To cancel the operation (display **[Lr]**). Press the **CAL** key and return to the weighing mode

Example of GX-AE/GX-A series

A-01111

The display shown left indicates the factory settings.

Function table

- 0 To inhibit changes to the function table.
- 1 To permit changes to the function table.

Calibration using the internal mass (One-touch calibration)

For balance software version 1.00 to 1.200

- 0 To inhibit calibration using the internal mass.
- 1 To permit calibration using the internal mass.

For balance software version 1.211 or later

- 0 *Lock* 0 To inhibit calibration using the internal mass.
- 0 *Lock* 1,2 To inhibit calibration using the internal mass when logged in as a user (*USER* 01 to 10) or a guest (*GUEST*). *1
- 1 To permit calibration using the internal mass.

Calibration using an external weight

For balance software version 1.00 to 1.200

- 0 To inhibit calibration using an external weight.
- 1 To permit calibration using an external weight.

For balance software version 1.211 or later

- 0 *Lock* 0 To inhibit calibration using an external weight.
- 0 *Lock* 1,2 To inhibit calibration using an external weight when logged in as a user (*USER* 01 to 10) or a guest (*GUEST*). *1

Automatic self calibration (Calibration due to changes in temperature)

- 0 To inhibit automatic self calibration.
- 1 To permit automatic self calibration.

Internal mass correction

- 0 To inhibit correction
- 1 To permit correction

*1 This is available when logged in as an administrator (*ADMIN*).

Example of GF-A series

A-00101

The display shown left indicates the factory settings.

Function table

0

To inhibit changes to the function table.

1

To permit changes to the function table.

No function

Calibration using an external weight

For balance software version 1.00 to 1.200

0

To inhibit calibration using an external weight.

1

To permit calibration using an external weight.

For balance software version 1.211 or later

0

Lock 0 To inhibit calibration using an external weight.

Lock 1,2 To inhibit calibration using an external weight when logged in as a user (USER 01 to 10) or a guest (GUEST). *1

No function

No function

*1 This is available when logged in as an administrator (ADMIN).

8-2 Initializing The Balance


This function returns the following parameters to factory settings. Calibration data

- Calibration data
- Function table
 - The sample unit mass value (counting mode),
 - 100% reference mass value (percent mode)
- The data that is stored in the balance using the data memory function
- External calibration weight and target weight value
- Function switch settings

Note Be sure to calibrate the balance (adjust sensitivity) after initialization.

Setting procedure

1. Press the **ON:OFF** key to turn off the display.
2. While pressing and holding the **PRINT** key and the **SAMPLE** key, press the **ON:OFF** key to display **P5**.
3. Press the **SAMPLE** key to display **[Lr]**.
4. Press the **PRINT** key.
To cancel this operation, press the **CAL** key.
5. Press the **RE-ZERO** key to change **No / 50**.
6. With displaying **[Lr 50]**, press the **PRINT** key to initialize the balance. The balance will automatically return to the weighing mode.









9. Function Table

The function table reads or rewrites the parameters that are stored in the balance. These parameters are maintained in non-volatile memory, even if the AC adapter is removed.










The function table menu consists of two layers. The first layer is the "Class" and the second layer is the "Item".

9-1 Setting The Function Table

Display symbol and keys


| | |
|---|--|
|  | The symbol "O" shows effective parameter. |
|  | When pressing and holding the key for 2 seconds in the weighing mode, the balance enters the function table mode. The key to select the class or item in the function table mode. |
|  | The key to change the parameter. |
|  | The key to change the parameter. |
|  | When a class is displayed, moves to an item in the class. When an item is displayed, stores the new parameter and displays the next class. |
|  | When an item is displayed, cancels the new parameter and displays the next class. When a class is displayed, exits the function table mode and returns to the weighing mode. |

Setting procedure

1. Press and hold the  key for 2 seconds until  of the function table is displayed in the weighing mode, then release the key.
2. Press the  key to select a class.
3. Press the  key to enter the class.
4. Press the  key to select a parameter for the selected item.
5. Press the  key to select an item.
6. To change another (multiple) item with the same class, repeat "4" and "5". To end the setting change of the same class, proceed to "7".
7. If storing parameters of the selected class, press the  key.
Then the next class is displayed.
If canceling the current operation, press the  key. Then the next class is displayed.
8. When specifying parameters for another class, proceed to "2".
When finishing the setting, press the  key to return to weighing mode.

Setting Example

This example sets "Stores weighing data (*DATA 2*)" for "Data memory (*DATA*)" and "1 minute (*INT 5*)" for "Interval time (*INT*)".



9-2 Details Of The Function Table

| Class | Item | Parameter | Description | |
|---|------------------------------|---------------------------------------|--|---|
| bRSFnC [00] Environment Display | Cond Condition | 0 | Fast response, sensitive value | Can be changed by response adjustment. With "Hold 1", sets the averaging time. |
| | | ▪ 1 | ↓ | |
| | | 2 | Slow response, stable value | |
| | St-b Stability band width | 0 | Stable when within ± 1 digit | The stabilization indicator illuminates with the display fluctuation within the range. With "Hold 1", sets the stabilization range. |
| | | ▪ 1 | ↓ | |
| | | 2 | Stable when within ± 3 digit | |
| | Hold Hold function | ▪ 0 | OFF | Holds the display when stable in animal mode. With "Hold 1", ANIMAL turns on. |
| | | 1 | ON | |
| | trc Zero tracking | 0 | OFF | Keeps zero display by tracking zero drift. |
| | | ▪ 1 | Normal | |
| | | 2 | Strong | |
| | | 3 | Very strong | |
| | SPd Display refresh rate | ▪ 0 | 5 times / second | Output frequency approx.5.21Hz |
| | | 1 | 10 times / second | Output frequency approx.10.42Hz |
| | | 2 | 20 times / second | Output frequency approx. 20.83hz |
| | Pnt Decimal point | ▪ 0 | Point (.) | Decimal point format |
| | | 1 | Comma (,) | |
| P-on Auto display-ON | ▪ 0 | OFF | Turns on the weighing mode display when AC adapter is connected | |
| | 1 | ON | | |
| P-off Auto display-OFF | ▪ 0 | OFF | Turns off the display after 10 minutes of inactivity. | |
| | 1 | ON | | |
| rnG Readability | ▪ 0 | Display readability | Display at weighing start | |
| | 1 | Not display readability | | |
| bEEP Buzzer | 0 | OFF | Buzzer sound such as key operation | |
| | ▪ 1 | ON | | |
| P-ZERO Stores tare value | ▪ 0 | OFF Zero indication at power on | ON Previous time weighing indication at power on | |
| | 1 | | | |
| d,SP-LED Backlight brightness | 0~9 | 10%~100% | | |
| | ▪ 5 | Factory setting 60% | | |
| LV-LED Bubble spirit level lightning | 0 | OFF | Bubble spirit level LED lightning | |
| | ▪ 1 | ON | | |
| ,SD Impact shock detection | 0 | OFF | Impact detection function (Valid for balance software version 1.300 or later) | |
| | ▪ 1 | ON | | |
| CL Add [01] Clock | | See "9-7 Clock And Calendar Function" | Confirms and sets the time and date. The time and date are added to output data. | |

▪ Factory setting

Note: "Digit" is a unit of readability.

- The number in [] is a class number. The numbers are shown as identifier when batching function settings together. Refer to "9-10 Output The Function Settings".

| Class | Item | Parameter | Description | | |
|--------------------------------------|-------------------------------------|----------------------------------|--|---|--|
| [02] [P Fnc] Comparator | [P Comparator mode | 0 | No comparison | | |
| | | 1 | Comparison when stable value or overloaded | | |
| | | 2 | Continuous comparison | | |
| | [P-t Number of comparator stages | 0 | 3 stage comparator | HI, OK, LO | |
| | | 1 | 5 stage comparator | HH, HI, OK, LO, LL | |
| | [P-z Near zero | 0 | Also compare near zero | | |
| | | 1 | ± 5 are not compared | | |
| | | 2 | ± 10 are not compared | | |
| | | 3 | ± 20 are not compared | | |
| | | 4 | ± 50 are not compared | | |
| | | 5 | ± 100 are not compared | | |
| | [P-p Polarity | 0 | Plus only | | |
| | | 1 | Minus only | | |
| | | 2 | Bipolarity | | |
| | [P in Input method | 0 | Digital input, upper / lower limits | [P HH, [P H ₁ , [P L ₀ | |
| | | 1 | Weighing input, upper / lower limits | [P LL can be selected. | |
| | | 2 | Digital input, reference value | [P rEF, [P LMt | |
| 3 | | Weighing input, reference value | [P LMt2 can be selected. | | |
| [P-Frd Flow measurement | 0 | Comparison by flow rate value | | | |
| | 1 | Comparison by weighing value (g) | | | |
| [03] [P VALUE Comparator value | [P HH Second upper limit | See "9-8 Comparator Function" | | Displayed only when [P in 0, 1 is set to digital input. | |
| | [P H ₁ Upper limit | | | | |
| | [P L ₀ Lower limit | | | | |
| | [P LL Second lower limit | | | | |
| | [P rEF Reference value | See "9-8 Comparator Function" | | Displayed only when [P in 2, 3 is set by input by load. | |
| | [P LMt Tolerance value | | | | |
| | [P LMt2 Second tolerance value | | | | |

■ Factory setting

Note: "Digit" is a unit of readability.

- The number in [] is a class number. The numbers are shown as identifier when batching function settings together. Refer to "9-10 Output The Function Settings".

| Class | Item | Parameter | Description | |
|---|-------------------------------|-----------|--|--|
| [04] CP bEEP Comparator buzzer | bEP HH HH buzzer | 0 | OFF | Displayed only when 5 step comparator is set. |
| | | 1 | ON | |
| | bEP HI HI buzzer | 0 | OFF | |
| | | 1 | ON | |
| | bEP OK OK buzzer | 0 | OFF | |
| 1 | | ON | | |
| bEP LO LO buzzer | 0 | OFF | | |
| | 1 | ON | | |
| bEP LL LL buzzer | 0 | OFF | Displayed only when 5 step comparator is set. | |
| | 1 | ON | | |
| [05] dout Data output | Prt Data output mode *1 | 0 | Key mode | Accepts the PRINT key only when the display is stable. |
| | | 1 | Auto print mode A (Reference=zero) | Outputs data when the weighing value stabilizes beyond the range from RP-P to RP-b from the zero point. |
| | | 2 | Auto print mode B (Reference=last stable value) | Outputs data when the weighing value stabilizes beyond the range from RP-P to RP-b from last stable value. |
| | | 3 | Stream mode | Outputs data at the specified display refresh rate. |
| | | 4 | Key mode B (Immediately) | Accepts the PRINT key regardless of the display condition. |
| | | 5 | Key mode C (When stable) | Accepts the PRINT key immediately when the display is stable, or waits for the display to be stable when not. |
| | | 6 | Interval output mode | Outputs data for each time set by int . |
| | RP-P Auto print polarity | 0 | Plus only | Displayed value > Reference |
| | | 1 | Minus only | Displayed value < Reference |
| | | 2 | Bipolarity | Regardless of displayed value |
| | RP-b Auto print difference | 0 | 10 digit | Difference between reference value and displayed value |
| | | 1 | 100 digit | |
| | | 2 | 1000 digit | |

■ Factory setting

Note: "Digit" is a unit of readability.

□ The number in [] is a class number. The numbers are shown as identifier when batching function settings together. Refer to "9-10 Output The Function Settings".

*1 download "Communication manual" from our website (<https://www.aandd.jp>) and refer to it.

| Class | Item | Parameter | Description | | |
|----------------------------------|---------------------------------|-----------|--|---|---|
| <i>dout</i> Data output mode | <i>dMEm</i> Data memory | 0 | OFF | Refer to "11. Data Memory" | |
| | | 1 | Stores unit mass in counting mode | | |
| | | 2 | Stores the weighing data and calibration history | | |
| | <i>int</i> Interval time | | 0 | Every measurement | Interval time in the interval memory mode when using <i>PrEtE</i> . |
| | | | 1 | 2 seconds | |
| | | | 2 | 5 seconds | |
| | | | 3 | 10 seconds | |
| | | | 4 | 30 seconds | |
| | | | 5 | 1 minutes | |
| | | | 6 | 2 minutes | |
| | | | 7 | 5 minutes | |
| | | | 8 | 10 minutes | |
| | <i>d-no</i> Data number | | 0 | No output | Valid when data memory function is ON. |
| | | | 1 | Output | |
| | <i>S-tD</i> Time/Date output | | 0 | No output | Refer to "9-7 Clock And Calendar Function" |
| | | | 1 | Time only | |
| | | | 2 | Date only | |
| | | | 3 | Time and date | |
| | <i>S-ID</i> ID number output | | 0 | No output ID number | |
| | | | 1 | Output ID number | |
| <i>PUSE</i> Data output pause | | 0 | OFF | Selects the data output interval. | |
| | | 1 | ON open 1.6 seconds | | |
| <i>At-F</i> Auto feed | | 0 | OFF | Selects whether or not automatic feed is performed. | |
| | | 1 | ON open 1 line | | |
| <i>inFo</i> GLP output | | 0 | OFF | Refer to "10-3 GLP Report" | |
| | | 1 | ON | | |
| | | 2 | ON (output clock of external) | | |
| <i>Ar-d</i> Zero after output | | 0 | OFF | Function to apply re-zero after outputting data. | |
| | | 1 | ON | | |
| <i>UFC</i> *1 UFC function | | 0 | OFF | Refer to "Communication manual" on the A&D website. | |
| | | 1 | ON | | |

■ Factory setting

Note: "Digit" is a unit of readability.

- The number in [] is a class number. The numbers are shown as identifier when batching function settings together. Refer to "9-10 Output The Function Settings".

*1 download "Communication manual" from our website (<https://www.aandd.jp>) and refer to it.

| Class | Item | Parameter | Description | |
|-------------------------------------|-----------------------------|------------------------|--|---|
| 5,IF [06] Serial interface *1 | Mode Access point | 0 | PC | All communication setting are possible. Only TYPE 0,1 can be selected. |
| | | 1 | Printer | |
| | | 2 | External indicator | |
| | bPS Baud rate | 0 | 600bps | |
| | | 1 | 1200bps | |
| | | 2 | 2400bps | |
| | | 3 | 4800bps | |
| | | 4 | 9600bps | |
| | | 5 | 19200bps | |
| | | 6 | 38400bps | |
| | bPr Data bit, parity bit | 0 | 7 bit EVEN | |
| | | 1 | 7 bit ODD | |
| | | 2 | 8 bit NONE | |
| | CrLF Terminator | 0 | CR LF | CR: ASCII 0Dh code LF : ASCII 0Ah code |
| | | 1 | CR | |
| | tYPE Data format | 0 | A&D standard format | Refer to "communication manual " on the A&D website. |
| | | 1 | DP format | |
| | | 2 | KF format | |
| | | 3 | MT format | |
| | | 4 | NU format | |
| 5 | | CSV format | | |
| t-UP Command time out | 0 | No limited | Selects wait time during command reception | |
| | 1 | Limited for one second | | |
| ErCd AK, error code | 0 | OFF | AK: ASCII 06h code | |
| | 1 | ON | | |
| USB [07] USB interface *1 | UFnc USB Function mode | 0 | Quick USB | Parameter will depend on the software version. |
| | | 1 | Bidirectional USB virtual COM | |
| | U-tP USB data format | 0 | A&D standard format | Refer to "communication manual " on the A&D website. |
| | | 1 | NU format | |
| | | 2 | CSV format | |
| | | 3 | TAB format | |
| | | 4 | NU2 format | |

■ Factory setting

Note: "Digit" is a unit of readability.

□ The number in [] is a class number. The numbers are shown as identifier when batching function settings together. Refer to "9-10 Output The Function Settings".

*1 download "Communication manual" from our website (<https://www.aandd.jp>) and refer to it.

| Class | Item | Parameter | Description | | |
|--|---|-----------|--|--|--|
| [10] <i>APP Fnc</i> Application function | <i>APP</i> Application mode | 0 | Normal weighing mode | Refer to "9-9 Description Of Application" | |
| | | 1 | Capacity indicator | | |
| | | 2 | Statistical calculation mode | | |
| | | 3 | Flow measurement mode | | |
| | | 4 | Gross, Net, Tare mode | | |
| | <i>STATF</i> Statistical function mode output items | 0 | Number of data, sum | | |
| | | 1 | Number of data, sum, max, min, range(max-min), average | | |
| | | 2 | Number of data, sum, max, min, range(max-min), average, standard deviation, coefficient of variation | | |
| | | 3 | Number of data, sum, max, min, range(max-min), average, standard deviation, coefficient of variation, relative error | | |
| | <i>FRD Unit</i> Flow rate unit | 0 | g / s (gram/second) | Refer to "13. Flow Measurement" | |
| | | 1 | g / m (gram/minute) | | |
| | | 2 | g / h (gram/hour) | | |
| | | 3 | mL / s (milliliter/second) | | |
| | | 4 | mL / m (milliliter/minute) | | |
| | | 5 | mL / h (milliliter/hour) | | |
| <i>Calc Auto</i> Calculation time automatic setting | 0 | OFF | | | |
| | 1 | ON | | | |
| [11] <i>MW Fnc</i> Minimum Weighing Warning Function | <i>MW-CP</i> Minimum weighing comparison | 0 | No comparison Do not use MW Fnc | Refer to "15. Minimum Weighing Warning Function" | |
| | | 1 | Comparison without near zero | | |
| | | 2 | Comparison including near zero | | |
| | <i>MW</i> Minimum weighing value input | | | | |
| | <i>MW-t</i> Minimum weight tolerance | 0 | 0.10% (standard deviation SDx2000) | | Refer to "15. Minimum Weighing Warning Function" |
| | | 1 | 1% (standard deviation SDx200 times) | | |
| <i>Minout</i> Minimum weight out | 0 | OFF | | | |
| | 1 | ON | | | |
| [12] <i>Unit</i> Unit | Refer to "4. Weighing" | | | | |
| [13] <i>dS Fnc</i> Specific gravity measuring function | <i>Ldin</i> Liquid density input | 0 | Water temperature | Refer to "18. Density (Gravity) Measurement" | |
| | | 1 | Liquid density | | |
| | <i>dS</i> Specific gravity measuring mode | 0 | Density measurement of a soil | | |
| | | 1 | Density measurement of a liquid | | |
| [14] <i>ALt</i> Programmable-unit (Multi-unit) | Sets an arbitrary coefficient. See "17. Programmable-Unit". | | Available only when programmable-unit mode is selected. | | |
| [15] <i>id</i> ID number setting | Refer to "10-2 Setting The ID Number" | | | | |

■ Factory setting

Note: "Digit" is a unit of readability.

- The number in [] is a class number. The numbers are shown as identifier when batching function settings together. Refer to "9-10 Output The Function Settings".

| Class | Item | Parameter | Description | |
|---|---|---|---|---------------------------------------|
| [16] <i>PASSwd</i> Password lock | <i>LOCK</i> Lock function | ▪ 0 | OFF | Refer to "19. Password Lock Function" |
| | | 1 | ON (Limit weighing operation) | |
| | | 2 | ON (Basic weighing is possible) | |
| | <i>PASSNo.</i> Password registration | <i>ADMIN</i> | Administrator password input | |
| | | <i>USER 01</i> | USER 1 password input | |
| | | <i>USER 10</i> | USER 10 password input | |
| [17] <i>AutoCAL</i> *2 Auto calibration | <i>TEMP</i> Calibration mode | ▪ 0 | Setting temperature | |
| | | 1 | Setting time | |
| | | 2 | Interval time | |
| | <i>TIME 1</i> Setting time1 | Refer to "7-1 Automatic Calibration (GX-AE/GX-A Series Only)" | | |
| | <i>TIME 2</i> Setting time2 | | | |
| <i>TIME 3</i> Setting time3 | | | | |
| <i>Int</i> Interval time | | | | |
| <i>IONFnc</i> *3 [20] Ionizer function | Refer to the instruction manual of "GXA-17 Large Glass Breeze Break With Ionizer" from our website. | | | |
| <i>CS in</i> *2 [18] Correction of internal mass value | <i>Auto</i> | Automatic input | Refer to "7-7-1 Correcting The Internal Mass Value Of The GX-AE/GX-A series (Auto)" | |
| | <i>MANUAL</i> | Digital input of correction value | Refer to "7-7-2 Correcting The Internal Mass Value Of The GX-AE/GX-A series (Manual)" | |

▪ Factory setting

Note: "Digit" is a unit of readability.

□ The number in [] is a class number. The numbers are shown as identifier when batching function settings together. Refer to "9-10 Output The Function Settings".

*2 GX-AE/GX-A series only.

*3 GX-AE series only.

9-3 Description Of The Class Environment Display

Condition ($[Cond]$)

$[Cond 0]$

This parameter is for sensitive response to the fluctuation of a mass value. Used for powder target mass, weighing a very light sample or when quick response weighing is required. After setting, the balance displays FAST.



$[Cond 2]$

This parameter is for stable weighing with slow response. Used to prevent a mass value from drifting due to vibration or drafts. After setting, the balance displays SLOW.

Stability band width ($[St-b]$)

This item controls the width to regard a mass value as a stable value. When the fluctuation per second is less than the parameter, the balance displays the stabilization indicator and outputs or stores the data by function setting (*double dRtR*, etc.) The parameter influences the "Auto print mode". Also, the readability being displayed is 1 digit.

Ex. If 0.01 g display is selected by pressing the SAMPLE key on the GX-303A, 0.01 g is 1 digit.

$[St-b 0]$

This parameter is used for sensitive response of the stabilization indicator. Used for exact weighing.



$[St-b 2]$

This parameter ignores slight fluctuations of a mass value. Used to prevent a mass value from drifting due to vibration or drafts

Hold function ($[Hold]$) (Animal weighing mode)

This function is used to weigh a moving object such as an animal. When the weighing data is over the weighing range from zero and the display fluctuation is within the stabilization range for a fixed period of averaging time, the processing indicator illuminates and the balance displays the average weight of the animal. When the animal or sample is removed from the weighing pan, the display returns to zero automatically. This function is available only when the hold function parameter is set to "1" (the animal mode indicator HOLD illuminates) and any weighing unit other than the counting mode is selected. The stabilization range and averaging time are set in "Condition ($[Cond]$)" and "Stability band width ($[St-b]$)".

| Weighing range | |
|----------------|-----------------|
| 0.0001g model | 0.0200g or more |
| 0.001g model | 0.200g or more |
| 0.01g model | 2.00g or more |
| 0.1g model | 20.0g or more |

| Averaging time | |
|----------------|-----------------------------|
| $[Cond 0]$ | 2 sec.(Efficiency priority) |
| $[Cond 1]$ | 4 sec. |
| $[Cond 2]$ | 8 sec.(Exact priority) |

| Stabilization range | | |
|---------------------|-----------|-------|
| $[St-b 0]$ | Lesser | 6.25% |
| $[St-b 1]$ | | 12.5% |
| $[St-b 2]$ | Greater * | 16.7% |

* Animal container kit (GXA-12) can be installed except GX-203A, GX-124A, GX-224A, GX-124AE, GX-224AE, GF-203A, GF-124A, GF-224A.

Zero tracking ($[trc]$)

This function tracks zero point drift caused by changes in the environment and stabilizes the zero point. When the weighing data is only a few digits, turn the function off for accurate weighing.

$[trc 0]$ The tracking function is not used. Used for weighing a very light sample.

$[trc 1]$ The normal tracking function is used. (± 1 digit / 1 second)

$[trc 2]$ The strong tracking function is used. (± 1 digit / 0.5 second)

$[trc 3]$ The very strong tracking function is used. (± 2 digit / 0.2 second)

Display refresh rate (5P_d)

The periodic time to refresh the display. This parameter influences "Baud rate", "Data output pause" and the data output rate of "Stream mode".

Decimal point (P_{nt})

The decimal point format can be selected.

Auto display-ON (P-_{on})

When the AC adapter is plugged in, the display is automatically turned on without the ON:OFF key operation, to display the weighing mode. Used when the balance is built into an automated system. Half an hour warm up (more than one hour for 0.0001g models) is necessary for accurate weighing.

Auto power-OFF (P-_{off})

This is a function to turn off only the display automatically when there is no operation made for a certain amount of time (approximately 10 minutes) while the power is on.

Readability (r_n_ū)

When weighing with rough precision, the readability can be turned off without key operation. This is useful when built into an automated system.

Buzzer (bEEP)

Select ON/OFF for the built-in buzzer that sounds when a key is operated or the status changes.

Tare value record (P-_{zEr})

After turning on the power supply, the display will not be automatically set to zero, and it will start from the previous weighing value. This is useful when a hopper, etc. is attached to the weighing pan and the power needs to be turned off while weighing discharge.

Backlight brightness (d_{5P}-LED)

Select the brightness of the backlight of the LCD display.

Bubble sprit level lightning (LV-LED)

Select ON/OFF for the LED that illuminates the bubble sprit level.

Impact shock detection (i_{5d})

Select ON / OFF for the function to display impact level.

* Applicable from balance software version 1.300 or later.

Even if the function to display impact level is turned off, it is recorded inside the balance when there is an impact.

9-4 Description Of The Data Output

Download "Communication manual" from our website (<https://www.aandd.jp>) and refer to it.

9-5 Description Of The Data Format

Download "Communication manual" from our website (<https://www.aandd.jp>) and refer to it.

9-6 Output Example Of The Data Format

Download "Communication manual" from our website (<https://www.aandd.jp>) and refer to it.

9-7 Clock And Calendar Function

The balance is equipped with a clock and calendar function. When the Clock and Calendar function (*dout, 5-td*) is set, the time and date are added to the output data.

Set or confirm the time and date as follows:

Operation

1. Press and hold the **[SAMPLE]** key for 2 seconds until **bASFnC** of the function table is displayed in the weighing mode, then release the key.
2. Press the **[SAMPLE]** key several times to display **[CLAdj]**.
3. Press the **[PRINT]** key.

The balance enters the mode to confirm or set the time and date.

Confirming the time


4. The current time is displayed with all the digits blinking.
 - When the time is not correct and is to be changed, press the **[RE-ZERO]** key and go to "5".
 - When the time is correct and the date is to be confirmed, press the **[SAMPLE]** key and go to "6".
 - When the time is correct and the date does not need to be confirmed, press the **[CAL]** key and go to "8".

Setting the time

5. Set the time in 24-hour format using the following keys.
 - [RE-ZERO]** (+) key..... To increase the value by one.
 - [MODE]** (-) key..... To decrease the value by one.
 - [SAMPLE]** key..... To select the digits to change the value.
The selected digits blink.
 - [PRINT]** key..... To store the new setting, display **[End]** and go to "6".
 - [CAL]** key..... To cancel the new setting and go to "6"

Confirming the date

6. The current date is displayed with all the digits blinking.
 - To change the display order of year (*Y*), month (*m*) and day (*d*), press the **[MODE]** key. The date is output in the order as specified.
 - When the date is not correct and is to be changed, press the **[RE-ZERO]** key and go to "7".
 - When the date is correct and the operation is to be finished, press the **[CAL]** key and go to "8".
 - When the time is to be confirmed again, press the **[SAMPLE]** key and go back to "4".



Setting the date

7. Set the date using the following keys. (The year is set with the last 2digits of the Christian era)

RE-ZERO(+)key ···· To increase the value by one.

MODE(-)key ······ To decrease the value by one.

SAMPLE key ······ To select the digits to change the value.
The selected digits blink.

PRINT key ········ To store the new setting, display **End** and go to "8".

CAL key ·········· To cancel the new setting and go to "8".

Quitting the operation

8. The balance displays the next menu (**PF Fnc**) of the function table. Press the **CAL** key to exit the clock and calendar function and return to the weighing mode.

Note Do not enter invalid values such as a non-existing date when setting the time and date. When the clock backup battery has been depleted, the balance displays **rtc PF**. Under this condition, press any key and set the time and date. The dead battery only affects the clock and calendar function. Even so, the function works normally as long as the AC adapter is connected to the balance.

9-8 Comparator Function

The comparison of comparators can select 3-steps or 5-steps ($[P F_{nc}]$, $[P-t]$) and it is set to 3-steps at the factory setting.

When 3-step comparator is set, the results of the comparison are indicated by $[HI]$ $[OK]$ $[LO]$ on the display.

When 5-step comparator is set, HH is indicated by $[HI]$ blinking and LL by $[LO]$ blinking.

By using GX-A-04, it is possible to output the comparison result at the contact point.

There are three types of scope that can be selected as follows.

- No comparison
- Comparison when the weight data is stable or overloaded
- Continuous comparison

The conditions for comparing near zero are in six levels from "including near zero" to "± 100 digits". "Upper limit value and lower limit value" and "reference value and tolerance range" are the comparison standards.

"Digital input" and "Input by sample load" are the for each value input methods for each value.

Refer to the function setting $[P F_{nc}]$.

By setting the function setting $[P bEEP]$, it is also possible to sound an internal buzzer depending on the result of the comparison.

3-step comparison result

| Weighing value | | 3-step comparison - display | | | |
|-----------------|--|-----------------------------|-------------|------------------|----------------|
| Threshold value | Judgment formula | Judgment result | Lit display | Blinking display | Buzzer control |
| Upper limit | Upper limit value < Weighing value | HI | $[HI]$ | | bEP H I |
| Lower limit | Lower limit value ≤ Weighing value ≤ Upper limit value | OK | $[OK]$ | | bEP oK |
| | Weighing value < Lower limit value | LO | $[LO]$ | | bEP L O |

5-step comparison result


| Weighing value | | 5-step comparison - display | | | |
|--------------------|--|-----------------------------|-------------|------------------|----------------|
| Threshold value | Judgment formula | Judgment result | Lit display | Blinking display | Buzzer control |
| Second upper limit | 2nd Upper limit value < Weighing value | HH | | $[HI]$ | bEP HH |
| Upper limit | Upper limit value < Weighing value ≤ 2nd Upper limit value | HI | $[HI]$ | | bEP H I |
| Lower limit | Lower limit value ≤ Weighing value ≤ Upper limit value | OK | $[OK]$ | | bEP oK |
| Second lower limit | 2nd Lower limit value ≤ Weighing value < Lower limit value | LO | $[LO]$ | | bEP L O |
| | Weighing value < 2nd Lower limit value | LL | | $[LO]$ | bEP LL |

Note

- The comparator function in the flow measurement mode ($[PPF]$) is compared at the factory setting with the flow rate value. By setting $[P-Frd]$ of the Function table $[P F_{nc}]$ to "1", it is also possible to compare with weight value (g unit).

Selecting comparators (3-steps and 5-steps)

1. Press and hold the **[SAMPLE]** key for 2 seconds until **bASFnC** of the function mode is displayed.
2. Press the **[SAMPLE]** key several times to display **[P Fnc]**.
3. Press the **[PRINT]** key.
4. Press the **[SAMPLE]** key several times to display **[P-t]**.
5. Press the **[RE-ZERO]** key to select "0" for 3 steps or "1" for 5 steps. Press the **[PRINT]** key to set.
6. Press the **[CAL]** key to return to the weighing mode.




Example1

(Always compare except "near zero" ± 20 digits and digitally input reference value and scope.)


Selecting a comparator mode (scope, comparison criteria and value for 3 step- comparator)

1. Press and hold the **SAMPLE** key for 2 seconds until **bR5FnC** of the function table is displayed, then release the key.
2. Press the **SAMPLE** key several times to display **[P Fnc]**.
3. Press the **PRINT** key.
4. Press the **RE-ZERO** key several times to display **[P ALL]** (" 2 " always compare).
5. Press the **SAMPLE** key several times to display **[P-]**.
6. Press the **RE-ZERO** key several times to display **[P-] E^x 20d** (" 3 " ± 20 digit is not compared.)
7. Press the **SAMPLE** key several times to move to **[P in]**.
8. Press the **RE-ZERO** key several times to display **[P in rFE DIG]**. (" 2 " reference value is set. digital input)
9. Press the **PRINT** key to store the selected mode.



Entering the values


10. With **[P VALUE]** displayed, press the **[PRINT]** key.
11. Display **[P rEF]**.
12. Press the **[PRINT]** key.
13. The current setting value is displayed with all of the digits blinking.
14. When the current setting is not to be changed, press the **[PRINT]** or **[CAL]** key to go to "15".



When the current setting is to be changed, press the **[RE-ZERO]** key and store the following keys.

- [SAMPLE]** key ····· Select the digit to change the value.
- [RE-ZERO]** key · Change the value of the digit selected.
- [MODE]** key ····· Switch the polarity.
- [PRINT]** key ····· Store the new setting and go to "15".
- [CAL]** key ······· Cancel the new setting and go to "15".

The current setting is not to be changed. The current setting is to be changed.



15. When **[P LME]** is displayed, pressing the **[PRINT]** key will display the currently set value.

If changing the setting value, it can be registered the tolerance value with the following keys.

For tolerance value, enter the value with the reference value set to 100%.

- [SAMPLE]** key ····· Move the blinking digit.
- [RE-ZERO]** (+)key · Change the value of the blinking digit.
- [MODE]** (-)key ····· Change the value of the blinking digit.
- [PRINT]** key ······· Register and go to "16"
- [CAL]** key ········· Cancel and go to "16"


16. Press the **[CAL]** key twice to return to the weighing display.

Example 2

(Continuous comparison, including "near zero", reference value and tolerance value.)


Selecting a comparator mode (scope, comparison criteria and value for 3-step comparator)

1. Press and hold the **SAMPLE** key for 2 seconds until **bASFnC** of the function table is displayed, then release the key.
2. Press the **SAMPLE** key several times to display **[P Fnc]**.
3. Press the **PRINT** key.
4. Press the **RE-ZERO** key several times to display **[P]** ("I" compared when stable and over)
5. Press the **SAMPLE** key several times to display **[P-]**.
6. Press the **RE-ZERO** key several times to display **[P-] IN**. ("0" near zero is also compared.)
7. Press the **SAMPLE** key several times to display **[P in]**.
8. Press the **RE-ZERO** key several times to display **[P in H/LWT]**. ("I" upper-lower limit is set. Input by loaded.)
9. Press the **PRINT** key to store the new setting.



Entering the reference and tolerance values

10. When **[CPVALUE]** is displayed, press the **[PRINT]** key. **[CPH_i]** will be displayed.
11. When **[CPH_i]** is displayed, press the **[PRINT]** key to check the currently set value (all blinking). Press the **[RE-ZERO]** key to enter the load input mode.
12. Press the **[RE-ZERO]** key, **[0.00 g]** is displayed. Place a sample of the weight of the upper limit on the balance and press the **[PRINT]** key. (Register the upper limit value.)
13. When finished, **[CPL_o]** is displayed. (Replace a sample of the weight of the upper limit from the balance.)
14. When **[CPL_o]** is displayed, press the **[PRINT]** key to check the currently set value (all blinking). Press the **[RE-ZERO]** key to enter the load input mode.
15. Press the **[RE-ZERO]** key, **[0.00 g]** is displayed.
16. Place a sample of the weight of the lower limit on the balance and press the **[PRINT]** key. (Register the lower limit value.)
17. When finished, **[CPH_i]** is displayed. (Replace a sample of the weight of the lower limit from the balance.)
18. Press the **[CAL]** key twice to return to the weighing display.



Sound the built-in buzzer corresponding to the comparison result.

1. Press and hold the **[SAMPLE]** key for 2 seconds until **bRSFnC** of the function table is displayed.
2. Press the **[SAMPLE]** key several times to display **CP bEEP**.
3. Press the **[PRINT]** key.
4. Press the **[SAMPLE]** key to set the buzzer sound setting of the comparison result "ON/OFF".

When 3-step comparator is set, the display can be selected from the following 3 kinds

bEP Hi, **bEP ok**, **bEP Lo**.*

When 5-step comparator is set, the display can be selected from the following 5 kinds


bEP HH, **bEP Hi**, **bEP ok**, **bEP Lo**

bEP LL.*

- [SAMPLE]** key Select the comparison result.
[RE-ZERO] key Set the buzzer sound setting for the comparison result to ON/OFF.
[PRINT] key Store the setting.

5. Press the **[PRINT]** key to set the buzzer sound setting for the comparison result. **End** is displayed and then **dout** is displayed.
6. Press the **[CAL]** key to return to weighing mode.

* Please refer to "Selecting comparators (3-steps and 5-steps)" on page 65 regarding about the comparator stage settings.



9-9 Description Of Application

Description of the normal weighing mode (*APF 0*).

The normal weighing mode of the factory setting.

Description of the weighing indicator mode (*APF 1*).

The weighing indicator displays the relation between load and weight in percent in normal weighing.

(Zero 0%, weighing 100%)

Note:

- It can not used with the settings " *dARA 1* or *dARA 2* " that use the data memory function (*dARA*).

Description of the statistical calculation mode (*APF 2*).

This is a function to statistically calculate the weighing value and to display and output the result.

Refer to "12. Statistical Calculation Mode " .

Description of the flow measurement mode (*APF 3*).

It is a function to calculate the flow measurement .

Refer to "13. Flow Measurement " .

Description of the gross net tare mode (*APF 4*).

This is a function to operate the setting and taring separately and to output the data of Gross (total amount), Net (net amount) and Tare (tare quantity).

Refer to "14. Gross Net Tare Function".

9-10 Output The Function Settings


With function settings, you can set the balance's operation to that appropriate for how it is used. In the menu structure of function settings, there are setting items in the classification item, and one setting value is registered in each setting item.

The status of function settings can be batch output and recorded by the following operation.

* This applies for balance software version 1.300 or later.

Procedure for batch output of function setting

1. Press and hold the **[SAMPLE]** key for 2 seconds in weighing mode.
2. **bR5Fnc** is displayed.
3. Press and hold the **[PRINT]** key for 2 seconds. **-List-** is displayed and current settings information is batch output.



Example of output

| | | |
|-------|------------|-----------------|
| | A & D | |
| MODEL | GX-324A | ← Type |
| S/N | 123456789 | ← Serial number |
| ID | LAB-0123 | ← ID |
| DATE | 2019/01/22 | ← Date |
| TIME | 16:29:35 | ← Time |

} Output date and time of the balance

Function Table

```

00,Cond ,1
00,St-b ,1
00,HoLd ,0
00,Trc ,1
00,SPd ,0
00,Pnt ,0
00,P-on ,0
00,P-oFF ,0
00,tnG ,0
00,bEEP ,0
00,P-ZEro ,0
00,diSP-LEd ,5
00,LV-LEd ,0
a b c
...
02,CP-Z ,2
02,CP-P ,2
02,CPin ,0
10,Frd Unit,0
10,Ct AUto ,0

11,MW-CP ,0
11,MW ,001.0000 g
11,MW-t ,0
11,Min out ,1

13,Ldin ,0
13,dS ,0

16,LocK ,0

17,CFnc ,0
17,Cint ,0

END
    
```

a. Classification number (2 digits)

b. Item (8 characters)

c. Parameter (1 or 12 digits)

* Each item is divided by a comma.

Please refer to "9-2 Details Of The Function Table" for the classification numbers, items and parameters.

Example 1 Output the function settings to a printer

Please use a multi-printer AD-8127 for printing.

1. Connect the balance and the printer.

When using the AD-8127, set the print mode to "DUMP". For details on setting and print mode, refer to the printer's instruction manual. For the connection of the balance and the printer, refer to "Communication Manual" from our website (<https://www.aandd.jp>).

2. Check that the balance and the printer can communicate, and output refer to "Procedure for batch output of function setting" in the previous section.

Example 2 Output the function settings to a computer

For more information about USB and WinCT settings, visit our website (<https://www.aandd.jp>) and refer to "Communication Manual" and "WinCT Operation Manual".

1. Connect the PC and balance with the supplied USB cable or RS-232C cable (sold separately).

* When using USB, please use in virtual COM mode. Output is not possible with Quick USB.

2. Install WinCT on your PC.

WinCT can be downloaded from our website (<https://www.aandd.jp>).

3. Start RSCom and match the communication settings such as COM port and baud rate with the balance.

When the [Start] button is pressed, communication becomes possible.

4. Check that the PC and the balance can communicate, and follow the previous section "Procedure for batch output of function setting" to output.

10.ID Number And GLP Report

10-1 Main Objective

- The data output compatible with "GLP/GMP" can be output to a personal computer or printer using the RS-232C serial interface.
- The GLP / GMP compliant report includes the balance manufacturer, model, serial number, ID number, date, time and space for signature. It includes the results and using mass for calibration or calibration test data.
- The balance can output the following GLP/GMP compliant reports from RS-232C or USB.
 - "Calibration report" of the calibration, using the internal mass (Calibration due to changes in auto and one-touch calibration.)
 - "Calibration report" of the calibration, using an external weight.
 - "Calibration test report" of the calibration test, using an external weight.
 - "Title block" and "End block" for the weighing data.
- Calibration report and calibration test data can be stored in memory to output several reports at the same time. Refer to "11. Data Memory" for details.
- The ID number is used to identify the balance when the balance is used for maintenance management.
- The ID number is maintained in non-volatile memory even if the AC adapter is removed.
- For details on confirming and setting the time and date, Refer to "9-7 Clock And Calendar Function".
- When printing the GLP output by connecting a multi-printer AD-8127 to the balance, the clock function of the printer can be used to print the time and date. (Function setting "FnF0 2") (Balance software version 1.211 or later)

This is enabled when centrally managing prevention of tampering of time and date with the password lock function on the AD-8127 side.

Note

When outputting GLP / GMP compliant reports, the print mode of AD-8127 is set to dump print mode. If the weighing value was printed in the external key printing mode, press and hold the **ENT** key on the AD-8127 for 2 seconds to switch between the external print mode and the dump print mode).


10-2 Setting The ID Number

1. Press and hold the **SAMPLE** key for 2 seconds until **bASFnC** of the function table is displayed, then release the key.
2. Press the **SAMPLE** key several times to display **id**.
3. Press the **PRINT** key. Set the ID number using the following keys.
 - SAMPLE** key..... To select the digit to change the value.
 - RE-ZERO** key, **MODE** key..... To set the character of the digit selected.
 - PRINT** key..... To store the new ID number and display **PASSwd**.
 - CAL** key..... To cancel the new ID number and display **PASSwd**.
4. With **PASSwd** displayed, press the **CAL** key to return to the weighing mode.

Note

The display segment of the balance is divided into 4 types.

For each segment display, refer to the "Display correspondence table" in the next page.



Display correspondence table

11 Segment

| | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 0 | 1 | 2 | 3 | 4 | 5 | 6 | 7 | 8 | 9 | - | □ | A | B | C | D | E | F | G | H | I | J | K | L | M | N | O | P | Q | R | S | T | U | V | W | X | Y | Z |
| 0 | 1 | 2 | 3 | 4 | 5 | 6 | 7 | 8 | 9 | - | □ | A | B | C | D | E | F | G | H | I | J | K | L | M | N | O | P | Q | R | S | T | U | V | W | X | Y | Z |

↳ Space

7 Segment

| | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 0 | 1 | 2 | 3 | 4 | 5 | 6 | 7 | 8 | 9 | - | □ | A | B | C | D | E | F | G | H | I | J | K | L | M | N | O | P | Q | R | S | T | U | V | W | X | Y | Z |
| 0 | 1 | 2 | 3 | 4 | 5 | 6 | 7 | 8 | 9 | - | □ | A | B | C | D | E | F | G | H | I | J | K | L | M | N | O | P | Q | R | S | T | U | V | W | X | Y | Z |

↳ Space

14 Segment

| | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 0 | 1 | 2 | 3 | 4 | 5 | 6 | 7 | 8 | 9 | - | □ | A | B | C | D | E | F | G | H | I | J | K | L | M | N | O | P | Q | R | S | T | U | V | W | X | Y | Z |
| 0 | 1 | 2 | 3 | 4 | 5 | 6 | 7 | 8 | 9 | - | □ | A | B | C | D | E | F | G | H | I | J | K | L | M | N | O | P | Q | R | S | T | U | V | W | X | Y | Z |

↳ Space

15 Segment

| | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 0 | 1 | 2 | 3 | 4 | 5 | 6 | 7 | 8 | 9 | - | □ | A | B | C | D | E | F | G | H | I | J | K | L | M | N | O | P | Q | R | S | T | U | V | W | X | Y | Z |
| 0 | 1 | 2 | 3 | 4 | 5 | 6 | 7 | 8 | 9 | - | □ | A | B | C | D | E | F | G | H | I | J | K | L | M | N | O | P | Q | R | S | T | U | V | W | X | Y | Z |

↳ Space

10-3 GLP Report

Set the function setting to "inf01" (use data of clock built in to the balance) or "inf02" (use clock data of external equipment) to output the GLP / GMP data with a AD-8127 (multi printer) or personal computer.


Note

- In case of outputting clock data built in the balance (inf01), if the time and date are not correct, set the correct time and date in "Clock (CLAdd)" of the function table.
- The setting of "inf02" can be set with the balance of software version 1.211 or later.

Calibration report using the internal mass

This is the GLP report when the balance is calibrated using the internal mass.

- Output the clock data of built in balance (*inF0 I*)




Output the clock data of external device (*inF0 2*).

By setting the function table "*inF0 2*" when outputting data such as GLP/GHP etc. It is possible to use the clock data of the external device such as PC or printer, not the built in balance data.

Note

- Clock data output from external device is for devices that have a clock function and can receive date and time data by receiving <ESC>D, <ESC>T.(Ex. AD-8127 multi printer, RsCom WinCT etc.)
- When saving the calibration history of the data memory function, the built in clock data is saved even if it is set to "*inF0 2*"



Calibration test report using the internal mass (0.0001g models only)

This is the GLP report when checking the weighing accuracy of the balance with the internal mass.

(Adjustment is not performed)

Setting of " *inF0!* "

Printer format (AD-8127)

```

                A & D
MODEL      GX-324A
S/N       123456789
ID        LAB-0123
DATE      2017/12/31
TIME      12:34:56
CAL.TEST<INT.>
ACTUAL
           0.0000 g
           +199.9999 g
TARGET
           +200.0000 g
REMARKS

SIGNATURE
-----
    
```

PC format (RsCom)

| | |
|----------------------|---------------------------|
| ← Manufacturer → | _____ A_&_D <TERM> |
| ← Model → | MODEL____GX-324A <TERM> |
| ← Serial number → | S/N____123456789 <TERM> |
| ← ID number → | ID_____LAB-0123 <TERM> |
| ← Date → | DATE__2017/12/31 <TERM> |
| ← Time → | TIME_____12:34:56 <TERM> |
| ← Calibration test → | CAL. TEST (INT.) <TERM> |
| | ACTUAL<TERM> |
| ← Zero point value → | _____ 0. 0000__g <TERM> |
| ← Target weight → | _____+199. 9999__g <TERM> |
| | value |
| ← Target weight → | _____+200. 0000__g <TERM> |
| | TARGET <TERM> |
| ← Remarks → | REMARKS <TERM> |
| | <TERM> |
| | <TERM> |
| ← Signature → | SIGNATURE <TERM> |
| | <TERM> |
| | <TERM> |

_____ Space, ASCII 20h
 <TERM> Terminator, CR LF or CR
 CR Carriage return, ASCII 0Dh
 LF Line feed, ASCII 0Ah

Calibration report using an external weight

This is the GLP report when the balance is calibrated using the external weight.

Setting of *inF0 l*

Printer format (AD-8127)

```

          A & D
MODEL   GX-10002A
S/N    123456789
ID     LAB-0123
DATE   2017/12/31
TIME   12:34:56
CALIBRATED(EXT.)
CAL.WEIGHT
      +10000.00 g
REMARKS

SIGNATURE
-----
  
```

— Space, ASCII 20h
 <TERM> Terminator, CR LF or CR
 CR Carriage return, ASCII 0Dh
 LF Line feed, ASCII 0Ah

PC format (RsCom)

```

          A_&_D <TERM>
MODEL_ GX-10002A <TERM>
S/N_ 123456789 <TERM>
ID_  LAB-0123 <TERM>
DATE_ 2017/12/31 <TERM>
TIME_ 12:34:56 <TERM>
CALIBRATED(EXT.) <TERM>
CAL.WEIGHT <TERM>
      +10000.00_g <TERM>
REMARKS <TERM>
<TERM>
<TERM>
SIGNATURE <TERM>
<TERM>
<TERM>
-----<TERM>
<TERM>
<TERM>
  
```

Calibration test report using an external weight

This is the GLP report when checking the weighing accuracy of the balance with the external weight.

(Adjustment is not performed)

Setting of *inF0 l*

Printer format (AD-8127)

```

          A & D
MODEL   GX-10002A
S/N    123456789
ID     LAB-0123
DATE   2017/12/31
TIME   12:34:56
CAL.TEST(EXT.)
ACTUAL
      0.00 g
      +9999.95 g
TARGET
      +10000.00 g
REMARKS

SIGNATURE
-----
  
```

— Space, ASCII 20h
 <TERM> Terminator, CR LF or CR
 CR Carriage return, ASCII 0Dh
 LF Line feed, ASCII 0Ah

PC format (RsCom)

```

          A_&_D <TERM>
MODEL_ GX-10002A <TERM>
S/N_ 123456789 <TERM>
ID_  LAB-0123 <TERM>
DATE_ 2017/12/31 <TERM>
TIME_ 12:34:56 <TERM>
CAL.TEST(EXT.) <TERM>
ACTUAL <TERM>
      0.00_g <TERM>
      +9999.95_g <TERM>
TARGET <TERM>
      +10000.00_g <TERM>
REMARKS <TERM>
<TERM>
<TERM>
SIGNATURE <TERM>
<TERM>
<TERM>
  
```

Heading and ending output

Application / Operation

As a method of managing weighing values, add "Heading" and "End" parts before and after the weighing value.


By pressing and holding the **PRINT** key for 2 seconds, "Heading" and "End" are output in turn.

Note

If the data memory function is used, heading and end cannot be output.

Key output method

1. While displaying the weighing value, hold down the **PRINT** key for 2 seconds and display **Start** to output "Heading".
2. Output the weighing value. The output method depends on the setting of the data output mode.
3. Press and hold the **PRINT** key for 2 seconds to display **Rec End**, "End" is output.
Setting of *Info!*



11.Data Memory

Data memory is a function to store weighing data and calibration data in memory. The data stored in memory are available for outputting at one time to a printer or personal computer.

The following six types of data can be stored.

| | |
|---|----------------|
| Unit mass (Counting mode) | Up to 50 sets |
| Weighing value | Up to 200 sets |
| Calibration report Internal calibration External calibration Calibration test report Internal test calibration (0.0001g models only) External test calibration | Last 50 sets |

11-1 Data Memory For Weighing Data

Features

- It is not necessary to connect the printer or personal computer to the balance continually, because the balance stores the weighing data in memory.
- By storing the weighing value in the balance, weighing operation can be performed without occupying the printer or PC for a long time.
- The data in memory can be displayed on the balance for confirmation.
- Data (ID number, data number, time and date) to be added to the output data can be selected in the function setting.
- The balance can store 200 sets of weighing data in memory (if time and date are added, the balance can store 100 sets).
 - * For the unit mass storage method, refer to "4-3 Counting Mode (PCS)".

Storing the weighing data


Note

1. Set the "Data memory (dAAR)" parameter to "dAAR 2". Refer to "9. Function Table".
2. Specify by the "Time/Date output (S-LD)" parameter whether time and date is to be added or not.
3. The storing mode depends on the "Data output mode (PrL)" parameter setting.
When set to PrL 3 (stream mode), data may not be stored correctly.


* It is also possible to change the time / date setting after storing the weighing value.

Enabling data memory function

1. Press and hold the **SAMPLE** key for 2 seconds until **bASFnC** of the function table is displayed.
2. Press the **SAMPLE** key several times to display **dout**.
3. Press the **PRINT** key.
4. Press the **SAMPLE** key several times to display **data**.
5. Press the **RE-ZERO** key to display **data Wt/CRl**.
6. Press the **PRINT** key to store the setting.
7. Press the **CAL** key to return to the weighing mode.



Display and symbol



When the volume of measured values stored reaches its maximum, **FUL** ↔ **data** blink in turn.

Caution

- When weighing data is being stored in memory, the data is output simultaneously using RS-232C interface or USB.
- "FUL" indicates that memory is full or the memory capacity has been reached. More data cannot be stored unless the memory data is deleted.
- Automatic self calibration can not be used while the interval memory mode is active.
- Statistic calculation function can not be used when the data memory function is active.

Setting the function table

Parameter settings for each output mode are as follows:

| Mode \ Item | Data output mode | Auto print polarity, difference | Data memory function | Interval time |
|------------------------|------------------|---------------------------------|----------------------|---------------|
| Key mode | <i>Pr</i> t 0 | Not used | <i>dAt</i> A 2 | Not used |
| Auto print mode A | <i>Pr</i> t 1 | <i>AP-A</i> 0 to 2 | <i>dAt</i> A 2 | |
| Auto print mode B | <i>Pr</i> t 2 | <i>AP-b</i> 0 to 2 | <i>dAt</i> A 2 | |
| Key mode B (immediate) | <i>Pr</i> t 4 | Not used | <i>dAt</i> A 2 | |
| Key mode C (stable) | <i>Pr</i> t 5 | | <i>dAt</i> A 2 | |
| Interval output mode | <i>Pr</i> t 6 | | <i>dAt</i> A 2 | |

Parameter settings for Data number, ID number, Time and Date


| | | | | | | |
|-------------|-----|-----------------|---------------|-----------|---------------|------------------|
| Data number | No | <i>d-no</i> "0" | Time and date | No | <i>S-td</i> 0 | — |
| | Yes | <i>d-no</i> "1" | | Time only | <i>S-td</i> 1 | Up to 200 pieces |
| ID number | No | <i>S-id</i> "0" | | Date only | <i>S-td</i> 2 | |
| | Yes | <i>S-id</i> "1" | | Both | <i>S-td</i> 3 | |

Recalling the memory data

Confirm that the "Data memory (*dAt*A)" parameter is set to "*dAt*A 2".

- Press and hold the **PRINT** key for 2 seconds until **RECALL** is displayed, then release the key.


The type of data appears in the upper left of the display as shown to the right "-d- or d-t".

Left of the display

 When setting without clock / date
 or

- Press the **PRINT** key to enter the memory recall mode.

Recall the data in memory using the following keys.

- RE-ZERO** key To proceed to the next data set.
- MODE** key To go back to the previous data set.
- PRINT** key To transmit the current data using the RS-232C or USB.
- CAL** key To exit the memory recall mode.


 When setting with clock / date


- Press the **CAL** key to return to the weighing mode.

* It is also possible to change the time / date output setting after storing the weighing value.

Transmitting all memory data at one time


Confirm that the "Serial interface (S,F)" parameters are set properly.
Refer to "9. Function Table" and "Communication Manual" on the A&D website (<https://www.aandd.jp>).

1. Press and hold the **PRINT** key for 2 seconds until **RECALL** is displayed, then release the key.
2. Press the **SAMPLE** key to display **out**.
3. Press the **PRINT** key to display **out No** with "No" blinking.
4. Press the **RE-ZERO** key to display **out 50** with "50" blinking.
5. Press the **PRINT** key to transmit all data using the RS-232C , USB.
6. The balance displays **CLEAR** when all data is transmitted.
Press the **CAL** key to return to the weighing mode.



Deleting all memory data at one time

1. Press and hold the **PRINT** key for 2 seconds until **RECALL** is displayed, then release the key.
2. Press the **SAMPLE** key several times to display **CLEAR**.
3. Press the **PRINT** key to display **CLEAR No** with "No" blinking.
4. Press the **RE-ZERO** key to display **CLEAR 50** with "50" blinking.
5. Press the **PRINT** key to delete all data
6. The balance displays **End** and returns to the weighing mode.




11-2 Data Memory For Calibration And Calibration Test

Characteristic

- Calibration data (when and how it is performed) and calibration test data can be stored in memory.
 - All the data in memory is available to be output at one time to a printer or personal computer.
 - Up to 50 data sets of the latest calibration or calibration test can be stored.
- * When the memory capacity has been reached to 50, "FUL" ↔ "CAL" illuminates in order in the upper left of the display as shown below.

Upper left of the display




Storing the calibration and calibration test data

1. Set the "Data memory (DATA)" parameter to "DATA 2". Refer to "9 Function Table".
2. With the settings above, each time calibration or calibration test is performed, the data is stored automatically.


Transmitting the memory data

1. Press and hold the **CAL** key for 2 seconds during weighing display. When a **CAL H.S** displayed, release your finger from the key to display **out**.
If there is no calibration history, **No dATA** is displayed, and then the display returns to the weighing display.
 2. Press the **PRINT** key to display **out No**.
 3. Change the **No** / **Go** with the **RE-ZERO** key. Display the **out Go**.
 4. Press the **PRINT** key to start output at one time while **out Go** is displayed.
The output format conforms to "GLP output".
 5. When output at one time is completed, **CLEAR** displays after **End** is displayed.
 6. If the saved history is deleted all at once, please proceed to "How to delete history". To return to the weighing value, press the **CAL** key.
- * If the **FUL** ↔ **CAL** indicators blink in turn during weighing display, 50 instance of data are stored. If history is saved history in this state, old data will be overwritten. Optionally delete the saved data.



Deleting data stored in memory

1. Press and hold the **CAL** key for 2 seconds until, **CAL H,5** is displayed, then release the key. **out** is displayed.
2. Press the **SAMPLE** key to display **CLEAR**.
3. Press the **PRINT** key to display **CLEAR No**.
4. Press the **RE-ZERO** key to change **No** / **Go**.
Display **CLEAR Go**.
5. Press **PRINT** while **CLEAR Go** is displayed, output at once is started.
6. When the balance displays **End** and returns to the weighing mode.



12. Statistical Calculation Mode

The statistical calculation mode statistically calculates the weight data, and displays or outputs the results. To use the statistical calculation mode, set the "Application function (APF)" parameter of "Application (AP Fnc)" in the function table to "2", as described below. To return to the normal weighing mode (factory setting), set "Application mode (APF)" to "0".

Statistical items available are number of data, sum, maximum, minimum, range (maximum-minimum), average, standard deviation and coefficient of variation. What statistical items to output can be selected from the four modes in the function table (SEPF).

- The wrong data input can be canceled by the key operation, if immediately after the input.
- Turning the balance off will delete the statistical data.
- The standard deviation and coefficient of variation are obtained by the equation below:

$$\text{Standard deviation} = \sqrt{\frac{N \cdot \sum (X_i)^2 - (\sum X_i)^2}{N(N-1)}} \quad \text{where } X_i \text{ is the } i\text{-th weight data,}$$

N is number of data.

$$\text{Coefficient of variation (CV)} = \frac{\text{Standard deviation}}{\text{Average}} \times 100 (\%)$$

$$\text{Relative error of maximum value} = \frac{\text{Maximum value} - \text{Average}}{\text{Average}} \times 100 (\%)$$

$$\text{Relative error of minimum value} = \frac{\text{Minimum value} - \text{Average}}{\text{Average}} \times 100 (\%)$$

Note

- When there is data with a readability digit off, the calculation result is displayed with the readability digit off. (Readability digit is rounded off.)
- When the data memory function is in use, the statistical calculation function cannot be used.
- When registering the warning function of the minimum weighing value, the statistical calculation function cannot be used.

12-1 How To Use The Statistical Calculation

Switching to the Statistical Function Mode (Changing The Function Table)


1. Press and hold the **SAMPLE** key for 2 seconds until **bASFnc** of the function table is displayed, then release the key.
2. Press the **SAMPLE** key several times to display **AP Fnc**.
3. Press the **PRINT** key to display **APF N=7**.
4. Press the **RE-ZERO** key several times to display **APF S=7t**.

To select statistical items to output, go to step 5.

To store the statistical function mode setting, go to 7.

To disable the statistical calculation mode, press the

RE-ZERO key to select **APF N=7**.



Selecting the statistical items to output

- Press the **SAMPLE** key to display $\overset{0}{\circ} \text{STAT} \text{ SUM}$.
- Press the **RE-ZERO** key to select the output items.
In the example, output the number of data, sum, maximum, minimum, range (maximum - minimum) and average are selected.

| Parameter | Description |
|-----------|---|
| 0 | Number of data, sum |
| 1 | Number of data, sum Maximum, minimum, range (maximum – minimum), average |
| 2 | Number of data, sum Maximum, minimum, range (maximum – minimum), average, Standard deviation, coefficient of variation |
| 3 | Number of data, sum Maximum, minimum, range (maximum – minimum), average, Standard deviation, coefficient of variation Relative error of maximum value, relative error of minimum value |

- Press the **PRINT** key to store the setting.
- Press the **CAL** key to return to the weighing mode.

Selecting the unit



- Press the **MODE** key to select the unit to be used for the statistical calculation mode. In the example shown at the right, gram (g) is selected.

Note

Selecting the unit using the **MODE** key is not available after the data is entered. In this case, clear the data as described on page 90 "Cleaning the statistical data" and select the unit using the **MODE** key.

When the unit used for the statistical calculation mode is to be enabled upon power-on, select the unit in "Unit (UNIT)" of the function table beforehand.

Continue from Step 4



Entering data for statistical calculation

Use the following keys to operate the statistical calculation mode.

MODE key When the data is entered, moves between the displaying items (weighing mode, statistical results and data operation) each time the key is pressed.

When no data has been entered, selects the unit.

SAMPLE key Turns the readability digit ON or OFF in the weighing mode.


RE-ZERO key Sets the display to zero in the weighing mode.

PRINT key Outputs the data number and the weight data and includes the weight data to statistical calculation in the weighing mode. (Output is not in the data format specified in the function table because of the data number added.)

Outputs the statistical results while the statistical results are displayed. (Output is not in the data format specified in the function table.)

CAL key Returns to the weighing mode.

1. Press the **RE-ZERO** key to set the display to zero.
2. Place the sample on the weighing pan and wait for the stabilization indicator to turn on.
3. Press the **PRINT** key to add the data displayed to statistical calculation. The number of data on the upper left of the display increases by 1.
4. Repeat steps 1 to 3 for each weighing.




Outputting the statistical results

- Each time the **MODE** key is pressed, the display changes: the results as selected in “Statistical function mode output items (SEAF)”, and **CLEAR**, **CANCEL**.

When pressing the **SAMPLE** key, the previous item is displayed.

Note

- When the number of data is 1, the coefficient of variation is displayed as **-----**.
 - When the average is 0, the coefficient of variation is displayed as **-----**.
 - Statistical items are indicated on the upper left of the display using the following symbols.
- When pressing the **PRINT** key while displaying the statistical result, the statistical result is output.




Output example

Function table parameter (SEAF)

| | |
|------|------------|
| N | 10 |
| SUM | + 1000.00g |
| MAX | + 105.00g |
| MIN | + 95.00g |
| R | + 10.00g |
| AVE | + 100.00g |
| SD | + 2.800g |
| CV | + 2.8 % |
| MAX% | + 5.0 % |
| MIN% | + 5.0 % |

| Symbol | Statistical item |
|--------|---------------------------------|
| Σuñ | Sum |
| ñRll | Maximum |
| ñln | Minimum |
| r | Range (Maximum – minimum) |
| RūE | Average |
| sd | Standard deviation |
| ēū | Coefficient of variation |
| ñRll% | Relative error of maximum value |
| ñln% | Relative error of minimum value |



Deleting the latest data


When the wrong data is entered, it can be deleted and excluded from statistical calculation. Only the latest data can be deleted.

- In the weighing mode, press the **MODE** key to display **CANCEL**.
- Press the **PRINT** key to display **CANCEL No.**
- Press the **RE-ZERO** key to display **CANCEL No.**
- Press the **PRINT** key to delete the latest data and exclude it from statistical calculation. The number of data decreases by 1 when the balance returns to the weighing mode.

Clearing the statistical data

All the statistical data will be deleted and the number of data will be 0 (zero).

1. In the weighing mode, press the **MODE** key.
2. The statistical data is displayed. Press the **MODE** key several times to display **CLEAR**.
3. Press the **PRINT** key to display **CLEAR No**.
4. Press the **RE-ZERO** key to display **CLEAR 0**.
5. Press the **PRINT** key to initialize the statistical data. The data count becomes 0 (zero) when the balance returns to the weighing mode.



12-2 Statistical Calculation Mode (Example Of Use)

Here, as an example of use of the statistical calculation mode, mixing of the multiple formula such as medicine is described. The mixing process is recorded using the balance and the printer.

In the example, the GX-303A and the AD-8127 are connected using the RS-232C serial interface.

Changing the function table

- Changes To enable the statistical calculation mode
 To enable “Zero after output”

Enabling the statistical calculation mode

1. Enter the function table menu.

Press and hold the **[SAMPLE]** key for 2 seconds until **bASFnC** of the function table is displayed, then release the key.

2. Select the application function.

Press the **[SAMPLE]** key several times to display **APFnC**. Then, press the **[PRINT]** key to display **APF Norm**.

3. Change the application function parameter to “2”.

Press the **[RE-ZERO]** key to display **APF² Stat**.

Press the **[PRINT]** key to confirm the change.

After **[End]**, **MWFnC** is displayed.

Enabling “Zero after output”

4. Select “Zero after output”.

Press the **[SAMPLE]** key several times to display **dout**.

Then, press the **[PRINT]** key to display **Pr⁰ KEI**,

and press the **[SAMPLE]** key several times to display

Pr⁰ -d OFF.


5. Enable “Zero after output”.

Press the **[RE-ZERO]** key to display **Pr¹ -d ON**. Then, press the **[PRINT]** key to confirm the change.

After **[End]**, **SIF** is displayed.

Returning to the weighing mode

6. Press the **[CAL]** key to return to the weighing mode.




Using The Statistical Calculation Mode

1. Press the **RE-ZERO** key to set the display to zero.
2. Place a container on the weighing pan.
Press the **PRINT** key to cancel the weight (tare). The balance displays **0.000 g**. (Storing the tare value)
The tare value data is output when the peripheral output equipment is connected.
3. Weigh formula 1 and press the **PRINT** key. The balance displays **0.000 g**. (Storing the weight value of formula 1)
The weight value data is output when the peripheral output equipment is connected.
4. Weigh formula 2 and press the **PRINT** key.
The balance displays **0.000 g**. (Storing the weight value of formula 2)
The weight value data is output when the peripheral output equipment is connected.
5. When there are some more formula to be added, repeat step 4.
6. After mixing is complete, press the **MODE** key to display the statistical results.
7. Press the **PRINT** key to output the number of data saved including the tare value and the total weight.

Output example

| | | |
|-------|-----------------|--------------------|
| No. 1 | ST, +0005.637 g | Tare value |
| No. 2 | ST, +0001.992 g | Formula 1 |
| No. 3 | ST, +0007.780 g | Formula 2 |
| N | 3 | |
| SUM | +15.409 g | Total weight |



13. Flow Measurement

The balance has a "flow mode" that calculates the amount of change in the weighing value per hour. For details, please refer to "FRD Addendum" which can be downloaded from the A&D website (<https://www.aandd.jp>).

- If the flow unit is set to mL/*, density can be registered. The maximum number of registrations is 10, and if density is set in advance, it can be selected according to the measurement sample.
- The flow rate value is calculated by the following formula.


$$Q = \frac{W - W'}{Ct}$$

Q : Flow rate

Ct : Calculation time

W : Current calculated value

W' : Weight value before Ct



For flow rate calculation time Ct, select manual / automatic and set.

13-1 How To Use Flow Measurement

Enable flow rate measurement

Switch flow rate measurement (Change the Function Table)


1. Press and hold the **SAMPLE** key for 2 seconds until **bASFnC** of the function table is displayed, then release the key.
2. Press the **SAMPLE** key several times to display **APFnC**.
3. Press the **PRINT** key to display **APF Norm**.
4. Press the **RE-ZERO** key several times to display **APF³ Frd**.

If you want to change the flow rate unit, go to 5.

If you want store the setting, go to 7.

If you want to cancel the flow function, press the

RE-ZERO key and return to **APF⁰ Norm**.



Setting of flow rate unit

5. Press the **SAMPLE** key to display $^{\circ} Frd Unit$.


6. Press the **RE-ZERO** key to set the setting value.

| Parameter | Contents |
|-----------|----------------------------|
| 0 | g / s (gram/second) |
| 1 | g / m (gram/minute) |
| 2 | g / h (gram/hour) |
| 3 | mL / s (milliliter/second) |
| 4 | mL / m (milliliter/minute) |
| 5 | mL / h (milliliter/hour) |

7. Press the **PRINT** key to store.

8. Press the **CAL** key to return to the calculating display.

- Factory setting



Manual / automatic selection of flow calculation time Ct.

There are two ways to set flow calculation time Ct, either by automatic setting in the balance according to the flow rate value or by manually determining a fixed value.

To switch between manual and automatic, please perform the following operation.


In factory setting, flow calculation time Ct is set to manual input setting ($Ct AUTO "OFF"$).

1. Please perform the following operation from the $^{\circ} Frd Unit mL/s$ display for flow unit setting as shown in "13-1 How To Use Flow Measurement".
2. Press the **SAMPLE** key to display $^{\circ} Ct AUTO$.
3. Press the **RE-ZERO** key to change ON/OFF.
4. Press the **PRINT** key to store.
5. Press the **CAL** key to return to the calculation display.

* If set to "OFF", refer to "How to set calculation time by manual setting" to set the flow calculation time.

If set to "ON", refer to "How to set calculation time by automatic setting" to set the flow calculation time.

From setting of flow rate unit



How to set flow calculation time by manual setting

The flow calculation time Ct can be set by the following procedure.

1. In weighing display, press and hold the **MODE** key for 2 seconds to display Ct 2 SEC.
2. Calculation time can be changed by following key operation.

The setting range is 1 second to 1 hour.


RE-ZERO (+) key ··· Change calculation time

MODE (-) key ····· Change calculation time

PRINT key ······· Store setting value

If the flow rate unit is g / * ,
the display will return to weighing display.
When the flow rate unit is mL / * ,
the display goes to density setting display.

CAL key ········· It returns to weighing display without storing the set value.



Note Unit of time setting (second(s), minute(m) or hour(h)) is entered in " * " of "g/*" and "mL/*".
For setting target values, refer to "GX-A/GF-A Series Flow Measurement Function Supplementary Manual".

How to set flow calculation time by automatic setting

It is possible to perform flow measurement without going to the trouble of selecting the flow rate calculation time Ct that matches the flow rate from the setting value.

The flow calculation time Ct is decided according to the flow rate value measured in 1 to 60 seconds. Accuracy can be selected from "Precision Priority (Resolution 500)", "Standard Setting (Resolution 200)" and "Response Priority (Resolution 50)".

The flow rate calculation precision can be changed by the following procedure.

1. Press and hold **MODE** key for 2 seconds to display **Fr RES** during weighing display.
2. Press the **RE-ZERO** key to change the desired setting value.

| Parameter | Description |
|-----------|-------------------------------------|
| 0 | Precision priority (Resolution 500) |
| 1 | Standard Setting (Resolution 200) |
| 2 | Response Priority (Resolution 50) |


■ Factory setting

3. Press the **PRINT** key to store.

If the flow rate unit is g / * , the display returns to weighing display or flow display.

If the flow rate unit is mL / * , the display transitions to density setting.
Please refer to "How to set the density".

Note Unit of time setting (second (s), minute (m) or hour (h)) is entered in " * " of "g/*" and "mL/*".



How to set the density

When the setting value of function setting **Frd Unit** is 3, 4, 5, after setting the calculation time, go to density setting display.

Density can be changed by following key operation.

The setting range is 0.0001g/cm³ to 9.9999g/cm³.

RE-ZERO(+) key Change the number of the blinking digit


MODE(-) key Change the number of the blinking digit

SAMPLE key Move the blinking digit

PRINT key The set value is the display returns to weighing display.

CAL key The display returns to weighing display without storing the set value.

From the Flow calculation time or Flow calculation precision



Method of reading density number

When flow unit is mL/*, up to 10 densities can be registered.

To register a new density, read the unconfigured density number and then register according to the procedure of the setting method of calculation time.

Continuing to hold down the **PRINT** key for 2 seconds in weighing display to display **d*.******.

Blinking **F**** is the current density number and **d*.****** is the set density value.

The density number can be changed by following key operation.

The setting range is F01 to F10.

Note

F** : The selected density number is entered.


d*.**** : The set density number is entered.

RE-ZERO(+) key Change density number.

MODE(-) key Change density number.

PRINT key Read the density of the selected density number and return to weighing display.

CAL key Return to the weighing display without reading the density of the selected density number.




Change display

After returning to the weighing value display after setting to flow mode, the unit is "g" with the **Frd** or **F**** indicator on.

Use the **MODE** key to switch between flow rate display and "g" display. By switching, the total amount and flow rate can be checked.

Note

F** : The selected density number is entered. (F01 to F10)



14. Gross Net Tare Function

Zero setting and taring can be operated separately, and data output for Gross (total amount), Net (net amount), Tare (tare quantity) becomes possible.

When the gross net tare function is selected, the key operation is changed as follows.

| Key | Operation |
|--------------------|---|
| ON:OFF key | Zero setting (Operate as the ZERO key) |
| RE-ZERO key | Tare (Operate as the TARE key) |

In order to use the Gross Net Tare Function, it is necessary to change the "setting of the Function table".


14-1 Preparation Of Gross Net Tare Function

To use this function, enter the Function table as follow, and set "Application Function $AP\ Fnc$ " to "4" in "Application mode APF ". To return the normal weighing mode (Factory setting), set "Application mode APF " to "0".

Please set as follows.



Setting procedure

1. Press and hold the **SAMPLE** key for 2 seconds until $bAS\ Fnc$ of the function table is displayed, then release the key.
2. Press the **SAMPLE** key several times to display $AP\ Fnc$.
3. Press the **PRINT** key to display $APF\ Norm$.
4. Press the **RE-ZERO** key several times to display $APF\ GNT$.
5. Press the **PRINT** key to store the setting.
6. Press the **CAL** key to return to the calculating display.



Key operation

In case of weighing value (gross) setting, operate with the following keys.

| Key | Function | Weighing value (gross) | Operation |
|---|---------------------|--|---|
|  | Zero setting (ZERO) | Within the zero range ^{*1} | Update a zero point and clear a tare value. |
| | | Out of the zero range ^{*1} | Do nothing |
|  | TARE | Plus value | Do tare and update a tare value |
| | | Gross zero ^{*2} (Gross zero mark blinking) | Clear a tare value |
| | | Minus value | Do nothing |

*1 "Zero range" means the range where the load is within $\pm 2\%$ of the weight from the reference zero.


For the zero range for each model, refer to "4-2 Basic Operation".

*2 "Gross zero" means the range where the readability of gross (total amount) is zero in "g".

(The state in which the gross zero mark is lit.)

Note To turn off the display of balance's display, press the ON:OFF key (Long press) for about 2 seconds.

Display




| No. | Mark | Description |
|-----|------|---|
| (1) | NET | This lights when the tare is not zero. |
| (2) | G | This lights when the tare is zero. |
| (3) | PT | When the preset tare is set by the PT command, this lights together with the NET mark. |
| (4) | gnt | This lights while using the gross net tare function. * Does not light with balance software version 1.300 or later. (When gross net tare function is in use, the NET or G mark lights always) |
| (5) | □ | This lights when the readability of the gross is in the range of zero in "g". |

Output

1. Every time pressing the **PRINT** key, it will output in the order of "NET" (object), "GROSS" (total amount), "TARE" (tare).
2. The compatible output format depends on the software version of the balance.

| Software version of the balance | Adaptive format |
|---------------------------------|---|
| 1.200 or later | A&D basic format DP format CSV format |

Output example (A&D standard format)




- By using the "UFC function", output connection and order also can be set.
 For the "UFC function", please refer to "Communication manual" which can be downloaded from the A&D website (<https://www.aandd.jp>).

14-2 Example Of Using The Gross Net Tare Function

1. After setting the gross net tare function, press the **ON:OFF** key when nothing is on the weighing pan. "G" will be displayed on the display.
2. Place the container to be tared on the weighing pan.
3. Press the **RE-ZERO** key to display **NET 0.00 g**, the tare value is set (updated). "NET" is displayed on the display.
4. Place the object.
5. Press the **PRINT** key, it will output in the order of "NET" (object), "GROSS" (total amount), "TARE" (tare).
6. Remove anything on the weighing pan and press the **ON:OFF** key to return to the "1".

To continue weighing without changing the tare value, remove the object only, place the next weighing object and press the **PRINT** key to continue outputting.



15. Minimum Weighing Warning Function

The minimum weighing value is the minimum necessary amount of sample to be used for correctly performing quantitative analysis, taking into consideration measurement error of the balance.

If the amount of sample is too small, the proportion of the measurement error in the measured value increases accordingly, and the reliability of the analysis result may drop.

By using the minimum weighing warning function, it is possible to judge at a glance whether the amount of sample meets the set minimum weight value. This function can be used only in "g" mode.

"*M,N*" is displayed at the top of the unit part when in use when the amount of sample is less than the set minimum weighing value, the "*M,N*" indication flashes.

When the amount of sample reaches the minimum weighing value or more, the "*M,N*" indicator will turn off.

The minimum weighing value can be changed from the function setting. The factory setting is 0 g.

If the set value is 0 g, no warning will be displayed even if the minimum weighing warning function is ON (*MW-CP* 1 or 2). Also, a value greater than weighing capacity cannot be set as minimum weighing value.

There are two kinds of warning display as follows

"Excluding near zero" *MW-CP* ± 0

"Including near zero" *MW-CP* ± 10


Near zero is within ± 10 digits of 0 g.

Note

- If *MW-CP* is set to anything other than *U*, the **MODE** key will be applied to set the minimum weighing value and units cannot be changed with the **MODE** key. (The latest unit will be fixed)
- To change the unit, turn OFF the minimum weighing value warning function.
- To turn OFF the minimum weighing value warning function, set *MW-CP* to *MW-CP* **OFF**, referring to "15-1 Minimum Weighing Comparison".

15-1 Minimum Weighing Comparison

1. Hold down the **SAMPLE** key to display the function setting **bASFnC**.
2. Press the **SAMPLE** key several times to display **MWFnC**.
3. Press the **PRINT** key.
4. **MW-CP** will be displayed. Press the **RE-ZERO** key to change the display from **MW-CP OFF** to **MW-CP E:0** (excluding near zero) or **MW-CP IN0** (including near zero).
5. If changing the setting of the minimum weighing value, proceed to 6. If not changing, press the **CAL** key to return to the weighing display.
6. Press the **SAMPLE** key to display **MW**.
(Make sure to check your balance software version and set the minimum weighing value.)




15-2 Input And Output Of The Minimum Weighing Value

15-2-1 Setting From The Function Setting (The Balance Software Version 1.200 To 1.220)

Input a minimum weighing value directly

Continues from step 6 of "15-1 Minimum Weighing Comparison".

7. When **MW** is displayed, press the **PRINT** key.
8. Set the minimum weighing value. The minimum weighing value can be changed by the following key operation.
 - RE-ZERO**(+) key ... Change the value of the blinking digit.
 - MODE**(-) key ... Change the value of the blinking digit.
 - SAMPLE** key ... Move the blinking digit.
 - PRINT** key ... Store the set value and proceed to the next item.
 - CAL** key ... Advance to the next item without storing the setting value.
9. Press the **PRINT** key to proceed to the next item.
10. Press the **CAL** key to return to the weighing display.



15-2-2 Setting From The Weighing Display (Balance Software Version 1.200 To 1.220)

1. Press the **MODE** key in the weighing display.
2. The current setting minimum weighing value blinks.
3. Press the **PRINT** key.
4. Set the minimum weighing value.

The minimum weighing value can be changed by the following key operation.


RE-ZERO(+) key ... Change the value of the blinking digit.

MODE(-) key ... Change the value of the blinking digit.

SAMPLE key ... Move the blinking digit.

PRINT key ... Store the set value and proceed to the next item.

CAL key ... Advance to the next item without storing the setting value.



15-2-3 Setting From The Function Setting (Balance Software Version 1.300 Or Later)

Input a minimum weighing value directly

Continues from step 6 of "15-1 Minimum Weighing Comparison".

7. When **MW** is displayed, press the **PRINT** key.
8. **KEY in** is displayed.

When setting a minimum weighing value:

Press the **PRINT** key again. Proceed to step 9.

When not setting a minimum weighing value:

Press the **CAL** key twice. The minimum weighing value will not be input, and the balance will return to weighing mode.

9. Set the minimum weighing value.

The minimum weighing value can be changed by the following key operation.

RE-ZERO(+) key ... Change the value of the blinking digit.

MODE(-) key ... Change the value of the blinking digit.


SAMPLE key ... Move the blinking digit.

PRINT key ... Store the set value and proceed to the next item.

- When **MW-CP** is set to **0**, the parameter is set to **1** automatically and the minimum value weighing comparison function will be available.

CAL key ... Advance to the next item without storing the setting value.

10. Press the **CAL** key to return to the weighing display.



Input with repeatability using external weight


Continues from step 6 of "15-1 Minimum Weighing Comparison".

7. When **MW** is displayed, press the **PRINT** key. **KEY in** is displayed.
8. Press the **SAMPLE** key. **EXT MASS** is displayed.
9. Press the **PRINT** key. **START**, **READY**, then the weighing display is displayed, and then the first load of repeatability will be requested with **LOAD** displayed.
10. Place the external weight on the balance. Then ◀ (in process mark) lights.
11. After ◀ (in process mark) blinks and the load is stable for 2 seconds, the span is displayed.
12. **REMOVE** blinks.
13. Remove the external weight. Then ◀ (in process mark) lights.
14. After ◀ (in process mark) blinks and the load is stable for 2 seconds, zero is displayed.
15. With **LOAD** displayed, the second load of repeatability will be requested. Subsequently, repeatability measurement will proceed until 10th time.
16. After the 10th span is displayed, **REMOVE**, **End**, then **MW^{0.1%} 23.20 g** of the minimum weighing display is displayed.


Error display description

- E g** The weight is too large.
- E g** The weight is too small.
- When the error is cleared, you are returned to repeatability measurement.
- Error!** When load is unstable for 20 seconds while repeatability is being measured, timeout occurs (more than 2 minutes passed without operations being made).
- After **Error** is displayed, the repeatability measurement shuts down and the balance will return to the function setting.

17. The repeatability (**5**) and the minimum weighing value (**MW**) can be output. When **MW^{0.1%} 23.20 g** is displayed, press the **SAMPLE** key to switch between minimum weighing value (**MW**) and repeatability (**5**). Press the **MODE** key to switch the measurement tolerance.




Ex: Batch output of minimum weighing values



18. Press the PRINT key to output the repeatability (SD) and the minimum weighing value (MW) selected at step 17.
Press and hold the PRINT key for 2 seconds to batch output the data.
19. When the output is completed, $MW\ 23.20\ g$ is displayed.
20. Press and hold the SAMPLE key for 2 seconds. The minimum weighing value is registered and you are returned to $MW\ 23.20\ g$.
- If $MW-CP$ is set to 0 , the parameter is automatically set to 1 and the minimum weighing comparison is available.
21. Press the CAL key twice to start the minimum weighing warning function by the weighing mode.

15-2-4 Setting From The Weighing Mode (Balance Software Version 1.300 Or Later)

1. Press the **MODE** key in the weighing mode.
2. When **MW 23.20^{KEY}g** is displayed, press the **PRINT** key.
3. **KEY in** is displayed.
After this, set the minimum weighing value setting from step 8 of "Input a minimum weighing value directly" or "Input repeatability using external weight" in "15-2-3 Setting From The Function Setting".



Displays the current set value above the unit

KEY Input set value.

EXT Input from repeatability using external weight.


ECL Input at ECL (press and hold the MODE key).

Refer to "6-2 Self Check Function / Automatic Setting Of Minimum Weighing Value By ECL".

15-2-5 Batch Output The Minimum Weighing Value (Balance Software Version 1.300 Or Later)

The result of setting the minimum weighing value and repeatability can be batch output.

1. Press the **PRINT** key for 2 seconds while **KEY in** or **EXT MASS** is displayed.
2. With the **REZERO** key, select "No" or "Go" from **out** display. By selecting "Go" and pressing the **PRINT** key, the result is batch output.
3. After the batch output is completed, **End** is displayed, and then the display returns to **KEY in** or **EXT MASS**.



Example: Batch output of the minimum weighing values set

The result depends on the settings.

Input with **KEY IN**

```
-MINIMUM WEIGHT-
                A & D
MODEL  GX-10002A
S/N    T2000112
ID     LAB-0123
DATE   2019/01/22
TIME   12:12:34

KEY INPUT

MINIMUM WEIGHT
        11.40 g

REMARKS

SIGNATURE
-----
```

Input with **EXT MASS**

```
-MINIMUM WEIGHT-
                A & D
MODEL  GX-10002A
S/N    T2000112
ID     LAB-0123
DATE   2019/01/22
TIME   12:34:56

EXTERNAL MASS

RESULT
1      +200.08 g
2      +200.07 g
3      +200.07 g
4      +200.07 g
5      +200.06 g
6      +200.07 g
7      +200.06 g
8      +200.07 g
9      +200.07 g
10     +200.07 g

SD      5.7 mg

TOLERANCE
        0.10 %
MINIMUM WEIGHT
        11.40 g

REMARKS

SIGNATURE
-----
```

Input with ECL

```
-MINIMUM WEIGHT-
                A & D
MODEL  GX-10002A
S/N    T2000112
ID     LAB-0123
DATE   2019/01/22
TIME   12:51:55

ECL

RESULT
1      +20.07 g
2      +20.06 g
3      +20.06 g
4      +20.06 g
5      +20.05 g
6      +20.06 g
7      +20.05 g
8      +20.06 g
9      +20.06 g
10     +20.06 g

SD      5.7 mg

TOLERANCE
        0.10 %
MINIMUM WEIGHT
        11.40 g

REMARKS

SIGNATURE
-----
```


15-3 Setting Measurement Tolerance Of Minimum Weighing Value

With the setting for Function Table $MW-L$, you can select the measurement tolerance of the minimum measured value (reference value) calculated from repeatability measurement by electronic control load (ECL). This function can be used for balance software version 1.200 to 1.220.

For balance software version 1.300 or later, please refer to "Input repeatability using external weight" in "15-2 Input And Output Of The Minimum Weighing Value" or "6-2 Self Check Function / Automatic Setting Of Minimum Weighing Value By ECL" for the setting.

Setting procedure

1. Press and hold the **SAMPLE** key for 2 seconds to display the function setting **bASFnC**.
2. Press the **SAMPLE** key several times to display **MWFnC**.
3. Press the **PRINT** key.
4. Press the **SAMPLE** key several times to display **MW-L**.
5. Press the **RE-ZERO** key to select $MW-L^0$ 0.10 % (Standard deviation SDx2000 times) or $MW-L^1$ 1 % (Standard deviation SDx200 times)
6. Press the **PRINT** key to move to the next item.
7. Press the **CAL** key to return to the weighing display.
8. For the repeatability measurement with setting value being by electronic control load (ECL), refer to "6-2 Self Check Function / Automatic Setting Of Minimum Weighing Value By ECL".




15-4 Data Output When Less Than Minimum Weighing Value

Data output ON/OFF can be switched with the setting for Function Table *Min out* when less than the minimum weighing value.

This function is available for balance software version 1.200 or later.

Setting procedure

1. Press and hold the **SAMPLE** key for 2 seconds to display the function setting **bRSFnc**.
2. Press the **SAMPLE** key several times to display **MW Fnc**.
3. Press the **PRINT** key.
4. Press the **SAMPLE** key several times to display **Min out**.
5. Press the **RE-ZERO** key to select **Min out ON**
(Data output ON) or **Min out OFF** (Data output OFF)
6. Press the **PRINT** key to move to the next item.
7. Press the **CAL** key to return to the weighing display.




16.Underhook

The underhook can be used for magnetic materials or density measurement. The built-in underhook is revealed by removing the plastic cap on the bottom of the balance.

Use the underhook as shown below.

Caution

- Do not apply excessive force to the underhook.
- When not in use, do not open the cover to prevent dust from getting into the balance.
- Do not push the underhook upward.



- The weighing pan, pan support and draft gate fall off, when turning over the balance. Remove them first.
- When not in use, attach the plastic cap to prevent dust from getting into the balance.

17. Programmable-Unit

This is a programmable unit conversion function. It multiplies the weighing data in grams by an arbitrary coefficient set in the function table and displays the result.

The coefficient must be within the range between the minimum and maximum shown below. The setting of coefficient differs according to the balance software version. If the coefficient set is beyond the range, an error is displayed and the balance returns to the coefficient setting mode, prompting to enter an appropriate value. A coefficient of 1 was set at the factory.

Balance software version 1.200 to 1.220

| Model | Minimum coefficient | Maximum coefficient |
|--|---------------------|---------------------|
| GF-123A GX/GF-203A/303A/403A/603A/1003A/1603A | 0.01 | 1000 |
| GF-1202A GX/GF-2002A/3002A/4002A/6002A/10002A | | 100 |
| GX/GF-6001A/10001A | | 10 |

Balance software version 1.300 or later

| Model | Minimum coefficient | Maximum coefficient |
|--|---------------------|---------------------|
| GX-124AE/224AE/324AE GX/GF-124A/224A/324A | 0.000001 | 10000 |
| GF-123A GX/GF-203A/303A/403A/603A/1003A/1603A | | 1000 |
| GF-1202A GX/GF-2002A/3002A/4002A/6002A/10002A | | 100 |
| GX/GF-6001A/10001A | | 10 |

Operation

1. Press and hold the **SAMPLE** key until **bASFnC** of the function table is displayed.
2. Press the **SAMPLE** key several times to display **MLE**.
3. Press the **PRINT** key. The balance enters the mode to confirm or set the coefficient.

Confirming the coefficient

4. The current coefficient is displayed with the first digit blinking.
 - When it is not to be changed, press the **CAL** key and proceed to step 6.
 - When it is to be changed, press the **RE-ZERO** key and proceed to step 5.


Setting the coefficient

5. Set the coefficient using the following keys.

SAMPLE key To select a digit to change the value. The selected digit blinks.


RE-ZERO key To change the value.

MODE key To change the decimal point position.
 Each time the switch is pressed, the decimal point position changes as follows:



PRINT key To store the new setting, display **End** and go to step 6.

CAL key To cancel the new setting and go to step 6.



Quitting the operation

6. The balance displays **Unit**. Press the **CAL** key to exit the programmable-unit function and return to the weighing mode.

Using the function

Press the **MODE** key to select the programmable-unit (no display on the unit section). Perform weighing as described in “4-2 Basic Operation (Gram Mode)”. After weighing, the balance displays the result (weighing data in grams x coefficient).

18. Density (Specific Gravity) Measurement

The balance is equipped with a density mode. It calculates the density of a solid using the mass value of a sample in air and the mass value in liquid.

For measurement, it use of the option GXA-13 specific gravity measurement kit is recommended.

Note

- The density mode was not selected for use when the balance was shipped from the factory. To use the mode, change the function table and activate the density mode "d5". Please refer to "Storing Units" of "4-1 Units".
- Readability is fixed when in density mode.

Formula to obtain the density

1. Density of solid

It can be obtained from the weight of the sample in air, the weight in the liquid, and the density of the liquid.

$$\rho = \frac{A}{A-B} \times \rho_0$$

ρ : Density of a sample A: Mass value of a sample in air
 ρ_0 : Density of a liquid B: Mass value of a sample in liquid

2. Density of liquid

Weight in air, weight in liquid and volume of float can be obtained using a float of a known.

$$\rho = \frac{A-B}{V}$$

ρ : Density of a sample A: Mass value of a sample in air
 V : Volume of float B: Mass value of a sample in liquid

(1) Prior to measurement: Changing the function table

Prior to measurement, change the function table as follows:

1. Register the density mode.

Density mode cannot be used at the factory setting.

Please refer to "Storing Units" of "4-1 Units" and register the gravimeter mode (d5).

Density mode is selected as one of the units with the MODE key.

2. Select whether the object to be measured is solid or liquid. (Function setting d5 Fnc, d5)

3. In the case of solid density measurement, select a method of inputting the density of liquid.

(function setting d5 Fnc, Ld in)

Density of liquid can be set by water temperature input or direct input of density, or input by the following function setting can be selected.

4. To start the measurement, display the weighing display.

Press the MODE key to display the specific gravity measurement display.

For the procedure, refer to (2) Method of measuring density (specific gravity) of solid or (4) Measuring the density of a liquid.

Note

- The following density function (d5 Fnc) is not displayed in the function settings unless density mode is enabled. First, perform the "Register the density mode" operation with the unit setting (Unit) of the function setting. When density mode is activated, "d5 Fnc" appears next to "Unit". For how to change the function setting, refer to "9. Function Table".

| Class | Item and parameter | Description |
|----------------------------|---------------------------------|--|
| d5 Fnc Density function | Ld in Liquid density input | ■ 0 Input water temperature |
| | | Input density directly |
| | d5 Measurement object select | ■ 0 Density measurement of solid |
| | | Density measurement of liquid |

■ Factory setting

(2) Method of measuring density (specific gravity) of solid (function setting $d5$ 0)

Note


- Re-set the density of the liquid with "(3) Entering the density of a liquid" as necessary, such as when the temperature of the liquid changes during measurement or when changing the type of liquid. In the density display, the 3 digits (4 digits for 0.0001g models) after the decimal point are fixed. The readability cannot be changed by pressing the **[SAMPLE]** key.

Density measurement displays the density fixed by measuring the weight in air and measuring the weight in liquid.

The relationship between each state and display is as follows.

Setting procedure

1. Check the weight measurement mode in air (g lights, blink \blacktriangleleft).
Press the **[RE-ZERO]** key to display zero without placing anything on the weighing pan.
2. Place the sample on the weighing pan in air and wait for the display to stabilize. If outputting the mass of the sample, press the **[PRINT]** key.
Next, press the **[SAMPLE]** key to fix the weight in air, and move to the weight measurement mode in liquid (g lights, blink \blacktriangleleft).
□ If auto-zero after data output ($\overline{A}r-d$) is set to on in the function setting, pressing the **[PRINT]** key to output will trigger the auto reset after output, preventing the density from being measured.
3. Transfer the sample from the weighing pan in air to the weighing pan in liquid and wait for the display to stabilize. If outputting the mass of the sample, press the **[PRINT]** key.
Next, press the **[SAMPLE]** key to fix the weight in liquid and shift to the density input mode (g turned off, lights \blacktriangleleft).
4. Enter the density of the liquid.
Refer to "(3) Entering the density of a liquid" and set the density.
Next, press the **[PRINT]** key to enter the density mode. (g turns off, lights \blacktriangleleft).
5. If outputting the density, press the **[PRINT]** key.
If measuring another sample, press the **[SAMPLE]** key and start with the weighing mode in air. The density unit is "g".
6. Re-set the density of the liquid with "(3) Entering the density of a liquid" as necessary, such as when the temperature of the liquid changes during measurement or when changing the type of liquid.
7. Press the **[MODE]** key to enter another weighing mode.




(3) Entering the density of a liquid

Two ways to set the density of a liquid are available in the function table, "Liquid density input (Ld_{in})" by entering the water temperature or by entering the density directly.

Entering the water temperature ($Ld_{in} T$)

The water temperature currently set (unit: °C, factory setting : 25°C) is displayed.



Use the following keys to change the value. Setting range is 0.0°C to 99.9°C, in increments of 0.1°C. Refer to the following matrix the " The relation between the water temperature and density".

RE-ZERO (+) key...The key to increase the temperature by one degree.
(0 is displayed after 9)

MODE (−) key The key to decrease the temperature by one degree.
(9 is displayed after 0)

SAMPLE keyMove the blinking digit.

PRINT keyThe key to store new water temperature and return to the density mode.(Proceed to Step 5)

CAL key The key to cancel the change and return to the density mode.
(Proceed to Step 5)


The relation between the water temperature and density

| °C | +0 | +1 | +2 | +3 | +4 | +5 | +6 | +7 | +8 | +9 |
|----|---------|---------|---------|---------|---------|---------|---------|---------|---------|---------|
| 0 | 0.99984 | 0.99990 | 0.99994 | 0.99996 | 0.99997 | 0.99996 | 0.99994 | 0.99990 | 0.99985 | 0.99978 |
| 10 | 0.99970 | 0.99961 | 0.99949 | 0.99938 | 0.99924 | 0.99910 | 0.99894 | 0.99877 | 0.99860 | 0.99841 |
| 20 | 0.99820 | 0.99799 | 0.99777 | 0.99754 | 0.99730 | 0.99704 | 0.99678 | 0.99651 | 0.99623 | 0.99594 |
| 30 | 0.99565 | 0.99534 | 0.99503 | 0.99470 | 0.99437 | 0.99403 | 0.99368 | 0.99333 | 0.99297 | 0.99259 |
| 40 | 0.99222 | 0.99183 | 0.99144 | 0.99104 | 0.99063 | 0.99021 | 0.98979 | 0.98936 | 0.98893 | 0.98849 |

g/cm³

Entering the density directly ($Ld_{in} I$)

The density currently set (unit : g / cm³, factory setting : 1.0000g / cm³) is displayed.



Use the following keys to change the value.

The range to set the density is 0.0000g / cm³ to 1.9999g / cm³.

If it is input beyond the settable range value, **Error2** is displayed and the display return to the input display.

RE-ZERO (+) key...The key to set the value of the blinking digit .(Next to 9 will be 0.)

MODE (−) key.....The key to select the blinking digit to change the value.(Next to 0 will be 9.)

SAMPLE keyMove the blinking digit.

PRINT keyThe key to store the change and return to the density mode. (Proceed to Step 5.)

CAL keyThe key to cancel the change and return to the density mode. (Proceed to Step 5.)

(4) Measuring the density of a liquid (Function table *d5 l*)


- In the density display, the 3 digits (4 digits for 0.0001g models) after the decimal point are fixed. Readability can not change with the **[SAMPLE]** key.

Density is displayed after "Mass measurement in air"
and "Mass measurement in liquid".

The procedure of each measurement is as follows:

Measuring procedure

1. Enter the density mode that "g (gram)" is displayed and the processing indicator (◀) blinks. Place nothing on both pan and press the **[RE-ZERO]** key to display zero.
2. Place the sample on the pan in air.
If the weight value is stored or output, press the **[PRINT]** key to store it after a stable weight value is displayed.
Press the **[SAMPLE]** key to decide the weight value in air and proceed to next step. (g lights, ▶ blinks)
- If auto-zero after data output (*R-r-d*) is set to on in the function setting, pressing the **[PRINT]** key to output will trigger the auto reset after output, preventing the density from being measured.
3. Place the liquid to measure the density of in the beaker and sink the float.
At this time, adjust so that the float is about 10 mm below the liquid level.
4. Move the sample to the pan in liquid.
If the weight value is stored or output, press the **[PRINT]** key to store it after a stable weight value is displayed. Press the **[SAMPLE]** key to decide the weight value in liquid and proceed to next step. (g turned off, *c-m* lights ▶ lights)
5. Enter the volume of the float.
Refer to "(5) Entering the volume of the float" and enter. Then press the **[PRINT]** key to return to the density mode.
6. If the density value is stored or output, press the **[PRINT]** key to store it. If the other sample is measured, press the **[SAMPLE]** key, and start from measurement of weighing mode in the air. The density unit is " *l* ".
7. Press the **[MODE]** key to proceed to other modes.




(5) Entering the volume of the float

The volume of the float that is currently set is displayed. (Factory setting is 10.00 c m³)

Change the setting value as follows.

The setting range is 0.01 cm³ to 99.99 cm³, every 0.01 cm³.



RE-ZERO (+) Key... The key to set the value of the blinking digit .
(Next to 9 will be 0.)

MODE (−) key The key to select the blinking digit to change the value.
(Next to 0 will be 9.)

SAMPLE key..... Move the blinking digit.

PRINT key The key to store the change and return to the density mode. (Proceed to Step 5.)

CAL key The key to cancel the change and return to the density mode. (Proceed to Step 5.)

19.Password Lock Function

By using the password lock function, it is possible to limit the usage and functions of the balance. This function is effective for preventing tampering of date / time setting and preventing internal setting changes by the user.

The password is set with four keys **MODE**, **SAMPLE**, **PRINT** and **RE-ZERO** keys in four digits (4 x 4 x 4 x 4 = 256 outcomes) .

At factory settings, the password lock function is disabled.

Enabling / disabling the password lock function and registering the password are performed in the function table.

Note

The function varies depending on the software version of the balance.

19-1 Balance Software Version 1.200

Two types of setting are possible depending on the "Lock" setting of the function table "Password lock function (PASSwd)".

| | |
|--------|---|
| Lock 0 | No password lock function |
| Lock 1 | Request password input at the start of weighing |

Lock 0 (No password lock function)

The password lock function is not be used.

Anyone can perform weighing work. In addition, all functions can be used and setting changes are also possible.

Lock 1 (Request password input at the start of weighing)

An administrator (AdMin.) can limit the users of the balance by setting individual passwords.

(The password input is required at the start of weighing with the **ON:OFF** key.)

The balance cannot be in weighing state unless you enter the correct password.

There are two login levels: Administrator (AdMin.) and user (OPRI. to 10)

| | |
|---------------------------|--|
| Administrator (AdMin.) | All functions and settings can be used. |
| | Passwords for 10 users can be set individually. |
| User (OPRI. to 10) | Initialization and setting changes are restricted (including clock). |
| No password | The balance cannot be used. |

Items that are limited by login level

| Login level | Weighing | | |
|---------------------------|----------------------------------|-------------|--------------------------------|
| | Password input at weighing start | Calibration | Change the function setting *4 |
| Administrator (AdMin.) | Possible | Possible | Possible |
| User (OPRI. to 10) | Possible | Impossible | Impossible |

*4 Changing response characteristics, setting minimum weight value, function selection and initialization and internal settings (setting of date and time).

19-2 Balance Software Version 1.211 Or Later

Three types of settings are possible depending on the "Lock" setting of the Function Table "Password lock function (PASSWORD)".

| | |
|--------|---|
| Lock 0 | No password lock function |
| Lock 1 | Request password input at the start of weighing |
| Lock 2 | To change the setting, login is required with the administrator's password. |

Lock 0 (No password lock function)

The password lock function is not used.

Anyone can perform weighing work. In addition, all functions can be used and setting changes are also possible.

Lock 1 (Request password input at the start of weighing)

An administrator (ADMIN) can limit the users of the balance by setting individual passwords.

(The password input is required at the start of weighing with the ON:OFF key.)

The balance cannot be in weighing state unless you enter the correct password.

There are two login levels: Administrator (ADMIN) and user (USER 01 to 10)

| | |
|-----------------------|--|
| Administrator (ADMIN) | All functions and settings can be used. Passwords for 10 users can be set individually. |
| User (USER 01 to 10) | Initialization and setting changes are restricted (including clock). |
| No password | The balance cannot be used. |

Lock 2 (To change the setting, login is required with the administrator's password.)

Anyone can perform weighing work, and initialization and setting changes can be restricted (including clock).

(Password input is not requested when weighing starts with the ON/OFF key.)

There are two levels of login level :Administrator (ADMIN) and user (GUEST)

| | |
|-----------------------|---|
| Administrator (ADMIN) | All functions and settings can be used. |
| Guest (GUEST) | Initialization and changing setting is restricted.(including clock) |

When weighing is started with the ON:OFF key while pressing the CAL key when the display is off, the password of the administrator (ADMIN) is requested.

Items that are limited by login level

| Login level | Weighing | | |
|-----------------------|----------------------------------|---------------|-------------------------------|
| | Password input at weighing start | Calibration | Change the function setting*5 |
| Administrator (ADMIN) | Necessary | Possible | Possible |
| User (USER 01 to 10) | | Impossible *6 | Impossible |
| Guest (GUEST) | Unnecessary | | |

*5 Changing response characteristics, setting of minimum weight value, confirming repeatability with built-in weight, function selection and initialization, internal settings (setting of date and time).

*6 The administrator (ADMIN) can set this to prohibited as shown in "8-1 Permit Or Inhibit".


19-3 Enable Password Lock Function

By the password function (*PASSwd*) of the Function table, the password function can be switched between "Invalid (*OFF*) / Valid (*ALL*) / Valid (*Fnc*)".

1. In the weighing mode, press and hold the **SAMPLE** key for 2 seconds to display **bASFnC**.
2. Press the **SAMPLE** key several times to display **PASSwd**.
3. Press the **PRINT** key to display **Lock OFF**.
(To cancel, press the **CAL** key.)
4. Press the **RE-ZERO** key to display **Lock ALL**.
(Press the **RE-ZERO** key again to display **Lock Fnc**.)
5. Press the **PRINT** key to display **SurE: YES No**.
("No" blinking while "No" selected.)
6. Press the **RE-ZERO** key to switch *YES* / *No*.
7. Display **SurE: YES No**. (*YES* blinking when selected *YES*.)
8. Press the **PRINT** key while *YES* is selected to enable the password lock function.
(With *Lock I*, password input is requested when the display is ON.)
9. **PASS No** is displayed. To register (change) the password, proceed to 4 on the "19-6 Registering Password (Changing)". If you will not register, press the **CAL** key twice to return to the weighing display.

Note

If the software version of the balance is 1.200, only switching between Invalid (*OFF*) / Valid (*ON*) can be selected.



19-4 How To Input The Password At The Start Of Weighing

Lock 1 (Request password input at the start of weighing)

1. Press the **ON:OFF** key while the display off.
2. After display **PASSword**, it becomes password input display **PH**.
3. Input 4 digits password using the following keys.
The balance will turn automatically after no operation for ten minutes.

MODE key Character **M**

SAMPLE key Character **S**

PRINT key Character **P**

RE-ZERO key Character **Z**


CAL key Back key

4. If the password is correct, the login level will be displayed, the weighing will be displayed after all lamps are illuminated. After entering the administrator's password, log in as an administrator.

(At factory settings, the password is set to

7777 by **RE-ZERO** key input 4 times at the administrator level.)


If the password is incorrect, the buzzer sounds 3 times, **FAIL** will display and the display will turn off.



Lock 2 (To change the setting, login is required with the administrator's password)

When logged in as a guest (**GUEST**)

1. Press the **ON:OFF** key while the display off.
2. After **LOG in: GUEST** displayed, return to the weighing display.




When logging in as administrator (ADMIN.) (Lock 1 or Lock 2)

1. Press the **ON:OFF** key with holding **CAL** key while the display off.
2. Input the 4 digits password using the following keys.
The balance will turn automatically after no operation for 10 minutes.

- MODE** key..... Character **M**
- SAMPLE** key..... Character **S**
- PRINT** key Character **P**
- RE-ZERO** key.... Character **Z**
- CAL** key Back key

3. If the password is correct, the login level will be displayed, the weighing will be displayed after all lamps are displayed.
(At factory settings, the password is set with **7777** of **RE-ZERO** key input 4 times at the administrator level.)
If the password is incorrect, the buzzer sounds 3 times in **FAIL** display and the display turns off.




Note

If the software version of the balance is 1.200, the login level at login is not displayed.
Also, for login at the administrator level, enter the administrator's password at weighing start with "Lock 1".

19-5 How To Logout

Log out by turning off the display by pressing the **ON:OFF** key.
If set to "Lock 1", the password will be requested again when switching the display from off to the weighing mode.



19-6 Registering Password (Changing)


The password can be changed at "Password (*PASS No.*)" of the Function Table.

1. Press and hold the **SAMPLE** key for 2 seconds in the weighing mode.
bRSFnc is displayed.
2. Press the **SAMPLE** key several times until **PASSwd** is displayed.
3. Press the **PRINT** key to display **Lock**.
4. Press the **SAMPLE** key to display **PASS No.**
5. Press the **PRINT** key to display **ADMIN.**
6. Press the **SAMPLE** key to change the login level **ADMIN. / USER 01 to 10**.
 If the password is already registered at the login level, the stability mark **o** is lit. (changeable)
7. Press the **PRINT** key to change the password.
 Refer to "19-7 Changing Password".

Note

- Log out by turning off the display to press the **ON:OFF** key.
- If set to **Lock 2**, the **ADMIN** password is required when logging in as an administrator.

Password registration of **USER 01 to 10** is unnecessary.



Refer to "19-7 Changing Password"

19-7 Changing Password

1. Refer to "19-6 Registering Password (Changing)" and display the login level you want to change the password on.
2. Press the **PRINT** key to display the current password. At factory settings, the password is **7777**. (the **RE-ZERO** key, 4 times)
3. Set the new password using the following keys. The balance will turn automatically after no operation for 10 minutes.

MODE key Character **M**

SAMPLE key Character **S**

PRINT key Character **P**

RE-ZERO key Character **Z**

CAL key Back key

CAL key(long press)..Delete password


(Refer to "19-8 How To Delete The Password *USER 01 to 10*)

4. Input 4 characters of the new password using these keys.
5. The balance displays **SurE: YES No** where "No" is blinking when "No" is selected. (Press the **CAL** key to return to the 4th character input.)
6. Press the **RE-ZERO** key to display **SurE: YES No** where "YES" is blinking when "YES" is selected.
7. Press the **PRINT** key to store the new password when "YES" is selected.
8. When the setting is completed, the next level is displayed. To continue the setting, set it from "6". To end the setting, press the **CAL** key 3 times to return to the weighing display.

Note

- If you forget your password, the balance can not be used. Please record and keep the password you registered.
- The password that is already registered by the administrator (**ADMIN**) cannot be registered by the user (**USER 01 to 10**).

Refer to "19-6 Registering Password (Changing)"




19-8 How To Delete The Password (USER 01 to 10)

1. Refer to "19-7 Changing Password" and select the user (USER 01 to 10) and display the Password input screen.
2. Hold down the **CAL** key when setting the password and display **CLEAR** (blink).
3. Press the **PRINT** key to display **CLEAR No**.
4. Press the **RE-ZERO** to change **Go / No**.
5. Press the **PRINT** when **CLEAR Go** to display **End** and delete the password.

Note

- The administrator's password cannot be deleted. Please refer to "19-6 Registering Password (Changing)" and "19-7 Changing Password" to change an arbitrary password.

Refer to "19-7 Changing Password"



19-9 Missing Password

If the correct password is missing, the balance can not be used.
Contact your local A&D dealer to reset the password to factory settings.

20.Repeatability Check Function (GX-AE/GX-A Series Only)

Repeatability is an indicator of variations in measured values when the same weight is repeatedly loaded and unloaded, and it is usually expressed in terms of standard deviation (σ_{n-1}).

The GX-AE/GX-A series have a built-in weight.

With the repeatability check function, the balance obtains 10 measurement data using the built-in weight and displays its standard deviation.

By installing the balance and using this function, it is possible to check repeatability in the environment where the balance is installed.


This function is available from balance software version 1.211 or later.

Ex. "Standard deviation = 10.0 mg" means that the result of repeated measurements of the same weighing material falls within the range ± 10.0 mg at a frequency of about 68%.

1. Press and hold the **[SAMPLE]** key for 4 seconds in the weighing display. After **rEP tEST** is displayed, release the key.
2. When **rEP tEST** is displayed, data collection starts automatically. While data is being collected, **rEP** blinks. **bASFnC** \leftrightarrow **rEP tEST**. To cancel, press the **[CAL]** key. **CANCEL** displays and you are returned to return to weighing mode.
3. When data collection is completed, repeatability (standard deviation) is displayed.
4. Press the **[PRINT]** key to output repeatability (standard deviation).
5. Press the **[CAL]** key to return to weighing mode.

Note

- The result of this function differs from the repeatability conditions of "26. Specifications" because it uses the weight of the balance (about 190g), so treat it as a reference value.
- This function cannot be used with GX-124A and GX-124AE.
- In order to measure correct data, do not apply wind or vibration while collecting data.
- While using the password lock function, it can be used only when logged in as **ADMIN** (administrator).



21. Interface Specification (Standard)

Download "Communication manual" from our website (<https://www.aandd.jp>) and refer to it.

22. Maintenance

22-1 Treatment Of The Balance

- Clean the balance with a lint free cloth that is moistened with warm water and a mild detergent.
- Do not use organic solvents to clean the balance.
- Do not disassemble the balance.
- Use the original packing material for transportation.

23. Troubleshooting

23-1 Checking The Balance Performance And Environment

The balance is a precision instrument. When the operating environment or the operating method is inadequate, correct weighing can not be performed. Place a sample on the pan and remove it, and repeat this several times. If the balance seems to have a problem with repeatability or to perform improperly, check as described below. If improper performance persists after checking, contact the local A&D dealer for repair.

"Frequently asked questions" and their answers are also posted on our website

<<https://www.aandd.jp>>.

1. Checking that the balance performs properly

- Please check the operation of the balance by the self diagnosis function.
Refer to "6-2 Self Check Function / Automatic Setting Of Minimum Weighing Value By ECL".
Fatal faults are indicated by messages.
- Check the balance performance using an external weight. Be sure to place the weight in the center of the weighing pan.
- Check the balance repeatability, linearity and calibrated value using external weights with a known value.

2. Checking that the operating environment or weighing method is proper

Operating environment

- Is the weighing table solid enough? (Especially 0.0001g and 0.001g model)
- Is the balance level? Refer to "2-2 Precaution" How to adjust the bubble spirit level.
- Is the operating environment free from vibration and drafts?
- Is there a strong electrical or magnetic noise source such as a motor near the balance?

Weighing method

- Does the weighing pan rim touch anything? Is the weighing pan assembly installed correctly?
- Is the RE-ZERO key pressed before placing a sample on the weighing pan?
- Is the sample placed in the center of the weighing pan?
- Has the balance been calibrated using the internal mass (one-touch calibration)? (Only for GX-AE/GX-A series)
- Has the balance been warmed up for one hour before weighing?

Sample and container















- Has the sample absorbed or lost moisture due to the ambient conditions such as temperature and humidity?
- Has the temperature of the container been allowed to equalize to the ambient temperature? Refer to "2-3 During Use".















Is the sample charged with static electricity? Refer to "2-3 During Use".

(This occurs especially with 0.0001g and 0.001g models when the relative humidity is low.)

- Is the sample of magnetic material such as iron? There are cautions about weighing magnetic materials. Refer to "2-3 During Use".


23-2 Error Codes

| Display | Error code | Description |
|---|------------|---|
|  | EC, E 1 1 | <p>Stability error</p> <p>The balance cannot stabilize due to an environmental problem. Check around the pan. Prevent vibration, drafts, temperature changes, static electricity and magnetic fields, from influencing the balance. Refer to "2-3 During Use".</p> <p>To return to the weighing mode, press the  key.</p> |
|  | | <p>Out of the setting range</p> <p>The data to be stored is out of the setting range.</p> |
|  | | <p>Malfunction of the internal memory element of the balance</p> <p>If this error continues to be displayed, repair is necessary. Please contact your dealer.</p> |
|  | EC, E 1 6 | <p>Internal mass error</p> <p>Applying the internal mass does not yield a change in the mass value as specified.</p> <p>Confirm that there is nothing on the pan and perform the weighing operation from the beginning again.</p> |
|  | EC, E 1 7 | <p>Internal mass error</p> <p>The internal mass application mechanism does not function properly. Perform the weighing operation from the beginning again.</p> |
|  | | <p>Abnormality in the internal memory data of the balance</p> <p>If this error continues to be displayed, repair is necessary. Please contact your dealer.</p> |
|  | | <p>Abnormality in the internal memory data of the balance</p> <p>If this error continues to be displayed, repair is necessary. Please contact your dealer.</p> |
|  | EC, E 2 0 | <p>Calibration weight error (Positive value)</p> <p>The calibration weight is too heavy. Confirm the calibration mass value. Press the  key to return to the weighing mode.</p> |
|  | EC, E 2 1 | <p>Calibration weight error (Negative value)</p> <p>The calibration weight is too light. Confirm the calibration mass value. Press the  key to return to the weighing mode.</p> |
|  | | <p>Overload error</p> <p>A sample beyond the balance weighing capacity has been placed on the pan. Remove the sample from the pan.</p> |
|  | | <p>Weighing pan Error</p> <p>The mass value is too light.</p> <p>Confirm that the weighing pan is properly installed and calibrate the balance.</p> |

| Display | Error code | Description |
|---|------------|---|
|  | | Sample mass error The balance can not store the sample for the counting mode or for the percent mode because it is too light. Use a larger sample. |
|    | | Unit mass error The sample unit mass for the counting mode is too light. Storing and using it for counting will cause a counting error. Add samples to reach the specified number and press the  key. Pressing the  key without adding samples will shift the balance to the counting mode. But, for accurate counting, be sure to add samples. |
|  | | Clock battery error The clock backup battery has been depleted. Press any key and set the time and date. The clock and calendar function works normally as long as the AC adapter is connected to the balance. If this error appears frequently, contact the local A&D dealer. |
|  | | Power supply voltage fault The voltage supplied from the AC adapter is abnormal. Please check if the problem is the AC adapter supplied with the balance. |
|   | | ECL repeatability With the self-check function, the standard deviation (SD) of repeatability due to electronically controlled load (ECL) exceeded 50 digits. Please revise the installation environment of the balance. <input type="checkbox"/>  This is displayed when repeatability is displayed by ECL. <input type="checkbox"/>  This is displayed when the minimum weighing value (reference value) by ECL is displayed. Refer to "6-2 Self Check Function / Automatic Setting Of Minimum Weighing Value by ELC." |
|  | | Full memory The maximum number or stored weighing values has been reached. In order to store more weighing values, it is necessary to delete the data. Refer to "11 Data Memory" |
|  | | Full memory The stored calibration and calibration test history has reached 50 instances. If more is stored, the old history will be deleted. Refer to "11 Data Memory" |
| | EC, E00 | Communications error A protocol error occurred in communications. Confirm the format, baud rate and parity. |
| | EC, E01 | Undefined command error An undefined command was received. Confirm the command. |

| Display | Error code | Description |
|---------|------------|--|
| | E C, E 0 2 | <p>Not ready</p> <p>A received command cannot be processed.</p> <p>Example: The balance received a "Q" command, but not in the weighing mode.</p> <p>The balance received a "Q" command while processing a RE-ZERO command.</p> <p>Adjust the delay time to transmit a command.</p> |
| | E C, E 0 3 | <p>Timeout error</p> <p>If the timeout parameter is set to "<i>t-UP</i>", the balance did not receive the next character of a command within the time limit of one second. Confirm the communication.</p> |
| | E C, E 0 4 | <p>Excess characters error</p> <p>The balance received excessive characters in a command. Confirm the command.</p> |
| | E C, E 0 6 | <p>Format error</p> <p>A command includes incorrect data.</p> <p>Example: The data is numerically incorrect.</p> <p>Confirm the command.</p> |
| | E C, E 0 7 | <p>Parameter setting error</p> <p>The received data exceeds the range that the balance can accept. Confirm the parameter range of the command.</p> |

23-3 Other Display



When this indicator (◀) blinks, automatic self calibration is waiting. If the balance is not used for several minutes with this indicator blinking, the balance automatically performs calibration using the internal mass. The blinking period depends on the operating environment.

Advice The balance can be used when this indicator is blinking, but we recommend that you perform calibration before weighing.

23-4 Asking For Repair

If the balance needs service or repair, contact your local A&D dealer.

The balance is a precision instrument. Use much care when handling the balance and observe the following when transporting the balance.

- Use the original packing material for transportation.
- Remove the weighing pan, pan support, breeze break ring and dust plate from the main unit.

24.Connection With Peripheral Device

Download "Communication manual" from our website (<https://www.aandd.jp>) and refer to it.

24-1 Command

Download "Communication manual" from our website (<https://www.aandd.jp>) and refer to it.

24-2 Key Lock Function


This function restricts the key operation of the balance by sending a specified command to the balance. Download "Communication manual" from our website (<https://www.aandd.jp>) and refer to it.

25.How To Check The Software Version Of The Balance

Specifications may vary depending on the software version of the balance.

1. Insert the AC adapter on the balance again.
2. The **LEVEL** display blinks.
3. After that, **P-*.***** will be displayed.

The number of " *.*** " becomes the software version.



26. Specifications

26-1 Common Models

26-1-1 Function

| | | |
|----------------------------------|---|---|
| Internal mass | For GX-A / GX-AE series (Approximately 190g) *1 | |
| Ionizer (static eliminator) | For GX-AE series | |
| Sensitivity drift (10°C to 30°C) | ±2ppm / °C (When auto calibration is OFF.) | |
| Operating environment | 5°C to 40°C, 85%RH or less (no condensation) | |
| Display refresh rate | 5 times/second, 10 times/second, or 20 times/second | |
| Display mode (unit) | g (gram), PCS (counting mode), % (percent mode), ct (Metric carat), mom (Momme) and density mode *2 | |
| Counting mode | Number of samples | 5, 10, 25, 50 or 100 pieces |
| Percent mode | Readability | 0.01%, 0.1%, 1% (Automatically changed by 100% mass) |
| Interface | RS-232C, USB | |
| AC adapter | Confirm that the adapter type is correct for the local voltage and power receptacle type Power consumption: Approx. 30 VA (supplied to the AC adapter) | |

*1 The mass of the internal mass may change due to corrosion or other damage caused by the operating environment, or it may change due to aging.

*2 For 0.0001g models, "mg" is available.

26-1-2 Dimension

| Model | 0.0001g | 0.001g | 0.01g | 0.1g |
|-----------------------|--|--|----------------------------|------|
| Weighing pan diameter | φ 90 mm | 128 x 128 mm | 165 x 165 mm | |
| Mass of product | Approx. 7 kg | Approx. 5 kg | Approx. 5 kg | |
| External dimensions | 259(w) x 358(D) x 332(H) mm (with glass breeze break) | 212(w) x 317(D) x 171(H) mm (with small breeze break) | 212(w) x 317(D) x 93(H) mm | |

26-2 Individual Models

26-2-1 0.0001g Models

| | | GX-124AE GX-124A GF-124A | GX-224AE GX-224A GF-224A | GX-324AE GX-324A GF-324A |
|--|-------------------|--|--|--|
| Weighing capacity | | 122g | 220g | 320g |
| Maximum display | | 122.0084g | 220.0084g | 320.0084g |
| Readability | | 0.0001g | | |
| Repeatability (Standard deviation) | | 0.0001g | | 0.0002g (300g) 0.0001g (200g) |
| Linearity | | ± 0.0002g | | ± 0.0003g |
| Stabilization time (FAST setting, good environment) | | Approx. 1.5 sec (100g) | Approx. 2 sec (200g) Approx. 1.5 sec (100g) | Approx. 2 sec (300g) Approx. 1.5 sec (100g) |
| Counting mode | Minimum unit mass | 0.0001g | | |
| Percent mode | Minimum 100% mass | 0.0100g | | |
| Carat | Weighing capacity | 610 ct | 1100 ct | 1600 ct |
| | Readability | 0.001 ct | | |
| Momme | Weighing capacity | 32 mom | 58 mom | 85 mom |
| | Readability | 0.0001 mom | | |
| External calibration weight | | 100g (factory setting) 100g 50g | 200g (factory setting) 200g 100g 50g | 200g (factory setting) 300g 200g 100g 50g |

26-2-2 0.001g Models

| | GF-123A | GX-203A GF-203A | GX-303A GF-303A | GX-403A GF-403A | GX-603A GF-603A | GX-1003A GF-1003A | GX-1603A GF-1603A | |
|---|---------------------------------------|---|---|---|---|---|--|---------|
| Weighing capacity | 122g | 220g | 320g | 420g | 620g | 1100g | 1620g | |
| Maximum display | 122.084g | 220.084g | 320.084g | 420.084g | 620.084g | 1100.084g | 1620.084g | |
| Readability | 0.001g | | | | | | | |
| Repeatability (Standard deviation) | 0.001g | | | | | | 0.002g (1600g) 0.001g (1000g) | |
| Linearity | ±0.002g | | | | | ±0.003g | | |
| Stabilization time (FAST setting, good environment) | Approx. 1 sec Approx. 0.8 sec (5g) | | | | | | Approx.1.5 sec Approx.0.8 sec (5g) | |
| Accuracy after calibration with internal mass * | ±0.010g | | | | | | ±0.010g (1000g) | |
| Counting mode | Minimum unit mass | 0.001g | | | | | | |
| Percent mode | Minimum 100% mass | 0.100g | | | | | | |
| Carat | Weighing capacity | 610 ct | 1100 ct | 1600 ct | 2100 ct | 3100 ct | 5500 ct | 8100 ct |
| | Readability | 0.005 ct | | | | | | |
| Momme | Weighing capacity | 32 mom | 58 mom | 85 mom | 112 mom | 165 mom | 293 mom | 432 mom |
| | Readability | 0.0005 mom | | | | | | |
| External calibration weight | 100g 50g | 200g (factory setting) 200g 100g 50g | 200g (factory setting) 300g to 100g (100g interval) 50g | 400g (factory setting) 400g to 100g (100g interval) 50g | 500g (factory setting) 600g to 100g (100g interval) 50g | 1000g (factory setting) 1000g to 100g (100g interval) 50g | 1000g (factory setting) 1600g to 100g (100g interval) 50g | |

* The operating environment does not include excessive change of ambient temperature, humidity, vibration, drafts, magnetic fields and static electricity.

26-2-3 0.01g Models

| | | GF-1202A | GX-2002A GF-2002A | GX-3002A GF-3002A | GX-4002A GF-4002A | GX-6002A GF-6002A | GX-10002A GF-10002A | |
|--|-------------------|--|--|--|--|--|--|--|
| Weighing capacity | | 1220g | 2200g | 3200g | 4200g | 6200g | 10200g | |
| Maximum display | | 1220.84g | 2200.84g | 3200.84g | 4200.84g | 6200.84g | 10200.84g | |
| Readability | | 0.01g | | | | | | |
| Repeatability (Standard deviation) | | 0.01g | | | | | 0.02g (10000g) 0.01g (5000g) | |
| Linearity | | ±0.02g | | | ±0.03g | | | |
| Stabilization time (FAST setting, good environment) | | Approx. 1 sec Approx. 0.8 sec (50g) | | | | | Approx. 1.5 sec (10g) Approx. 0.8 sec (50g) | |
| Accuracy after calibration with internal mass * | | ±0.10g | | ±0.15g | | ±0.15g (5000g) | | |
| Counting mode | Minimum unit mass | 0.01g | | | | | | |
| Percent mode | Minimum 100% mass | 1.00g | | | | | | |
| Carat | Weighing capacity | 6100 ct | 11000 ct | 16000 ct | 21000 ct | 31000 ct | 51000 ct | |
| | Readability | 0.05 ct | | | | | | |
| Momme | Weighing capacity | 325 mom | 586 mom | 853 mom | 1120 mom | 1653 mom | 2720 mom | |
| | Readability | 0.005 mom | | | | | | |
| External calibration weight | | 1000g 500g | 2000g (factory setting) 2000g 1000g 500g | 2000g (factory setting) 3000g to 1000g (1000g interval) 500g | 4000g (factory setting) 4000g to 1000g (1000g interval) 500g | 5000g (factory setting) 6000g to 1000g (1000g interval) 500g | 10000g (factory setting) 10000g to 1000g (1000g interval) 500g | |

* The operating environment does not include excessive change of ambient temperature, humidity, vibration, drafts, magnetic fields and static electricity.

26-2-4 0.1g Model

| | | GX-6001A GF-6001A | GX-10001A GF-10001A |
|--|-------------------|--|--|
| Weighing capacity | | 6200g | 10200g |
| Maximum display | | 6208.4g | 10208.4g |
| Readability | | 0.1g | |
| Repeatability (Standard deviation) | | 0.1g | |
| Linearity | | ±0.1g | |
| Stabilization time (FAST setting, good environment) | | Approx. 1 sec Approx. 0.8 sec (500g) | |
| Accuracy after calibration with internal mass * | | ±0.5g (5000g) | |
| Counting mode | Minimum unit mass | 0.1g | |
| Percent mode | Minimum 100% mass | 10.0g | |
| Carat | Weighing capacity | 31000 ct | 51000 ct |
| | Readability | 0.5 ct | |
| Momme | Weighing capacity | 1653 mom | 2720 mom |
| | Readability | 0.05 mom | |
| External calibration weight | | 5000g (factory setting) 6000g to 1000g (1000g interval) 500g | 10000g (factory setting) 10000g to 1000g (1000g interval) 500g |






* The operating environment does not include excessive change of ambient temperature, humidity, vibration, drafts, magnetic fields and static electricity.

27. External Dimensions

GX-124AE / GX-224AE / GX-324AE


GX-124A / GX-224A / GX-324A

GF-124A / GF-224A / GF-324A




- *1 Opening height of side sliding door
- *2 Maximum width when side sliding door is open
- *3 Inside dimension
- *4 Height of weighing pan
- *5 Height from the weighing pan to the top of the sliding door of the glass breeze break
- *6 Area under the floor weighing platform
- *7 DC jack protruding dimension of AC adapter
- *8 Maximum height of the sliding door when open
- *9 Width of the top of the sliding door when open
- *10 Can not open doors on both sides at the same time.

GX-203A / GX-303A / GX-403A / GX-603A / GX-1003A / GX-1603A
 GF-123A / GF-203A / GF-303A / GF-403A / GF-603A / GF-1003A / GF-1603A



- *1 Opening width when transparent plate is removed.
- *2 Inside dimension
- *3 Weighing pan size
- *4 Height from the weighing pan to the lid of the breeze break.
- *5 Opening height when transparent plate is removed.
- *6 Area under the floor weighing platform.
- *7 DC jack protruding dimension of AC adapter.

GX-2002A / GX-3002A / GX-4002A / GX-6002A / GX-10002A/ GX-6001A/ GX-10001A
 GF-1202A / GF-2002A / GF-3002A / GF-4002A / GF-6002A / GF-10002A/ GF-6001A /
 GF-10001A



- *3 Weighing pan size
- *6 Area under the floor weighing platform.
- *7 DC jack protruding dimension of AC adapter.

27-1 Options And Peripheral Instruments

Options

GXA-03 : 2nd RS-232C interface

- RS-232C Interface insulation type for expansion

GXA-04 : Comparator relay output / buzzer / external key input intercafe

- Outputs comparator results.

GXA-06 : Analog output interface (factory installed / dealer option)

- This option outputs a voltage of 0 to1V (or 0.2 to1V).

Fxi-08 : Ethernet interface

- Enables the balance to communicate with computers on a network.
- Multiple balances on a network can be controlled by one computer.

GXA-09 : Built-in rechargeable battery (factory-installed / dealer option)

- Enables the balance to be used in an environment where the AC adapter can not be used.
- Charging time is approximately 10 hours, and continuous use time is approximately 14 hours.

GXA-10 : Large glass breeze break

- Breeze break unit with a glass door

GXA-12 : Animal weighing pan (for models of 320g capacity or higher)

- Container with depth to make it difficult for animals to escape

GXA-13 : Density determination kit (for the 0.001g models only)

- Unit that enables easy weighing of the sample's weight in air and in water.

GXA-14 : Density determination kit (for the 0.0001g models only)

- Unit that enables easy weighing of the sample's weight in air and in water.

GXA-17 : Large glass breeze break with built-in fanless ionizer and external IR switch

- GX-10 breeze break unit with ionizer

GXA-23-PRINT : External key input interface + the AX-SW137-PRINT foot switch

- External contact input terminal that can operate PRINT and RE-ZERO key.
Foot switch of print function included. (AX-SW137-PRINT)

GXA-23-RE-ZERO : External key input interface + the AX-SW137-REZERO foot switch

- External contact input terminal that can operate PRINT and RE-ZERO key.
Foot switch of RE-ZERO function included. (AX-SW137-REZERO)

GXA-23-PLUG : External key input interface + the AX-T-314A-S plug

- External contact input terminal that can operate PRINT and RE-ZERO key.
Three assembled stereo plugs are included.

Note: In order to use, it is necessary to solder the attached plug and the switch prepared by the customer.

GXA-24 : USB host interface (factory-installed / dealer option)

- Stores the weighing value in the USB memory.

GXA-25 : Fanless ionizer of Quick Ion technology

- This ionizer (static eliminator) unit can be connected to and installed near the GX-AE/GX-A/GF-A series.

GXA-26 : External IR switch

- External touchless switch that can operate PRINT and RE-ZERO key.

AX-GXA-31 : Main unit cover (5 pieces)

- Protective cover for standard accessories

Peripheral devices

AD-8920A : Remote Display

- This option can be connected to the balance using the RS-232C interface or current loop and displays the weighing data transmitted by the balance.

AD-8922A : Remote Controller

- This option can be connected to the balance using the RS-232C interface and can control the balance remotely.

AD-8127 : Compact printer

- Small dot impact printer that connects with the balance via the RS-232C interface.
- Statistical function, clock and calendar function, interval print function, graphic print function, dump print mode

AD-1687 : Weighing Environment logger

- A data logger equipped with 4 sensors for temperature, humidity, barometric pressure and vibration that can measure and store environmental data. When connected to the RS-232C interface of the balance, the AD-1687 can store environmental data along with weighing data. Therefore, it is possible to store data in an environment where a computer can not be used.

AD-1688 : Weighing data logger

- When connected to the RS-232C interface of the balance, the AD-1688 can store the data in an environment where a personal computer can not be used.

AD-1689 : Tweezers for calibration weight

- A pair of tweezers ideally suited for holding calibration weights of 1g to 500g.

AX-USB-9P : Serial / USB Converter

- An RS-232C cable is provided to connect the USB converter to the balance.
- Enables bi-directional communication between the PC and the balance when a USB driver is installed.

AX-SW137-PRINT : Foot switch for print (with connector)

- Foot switch that functions in the same way as the PRINT key when combined with GXA-23 external connector

AX-SW137-REZERO : Foot switch for re-zero (with connector)

- Foot switch that functions in the same way as the RE-ZERO key when combined with GXA-23 external connector

AX-BM-NEEDLESET : Electrode units for the ionizer (4 pieces)

- Electrode unit replacement for ionizer. When replacing, please replace two at the same time. For how to replace, refer to "Maintenance of the electrode unit" in the manual of "GXA-17 Large Glass Breeze Break with Ionizer" which can be downloaded from our website (<https://www.aandd.jp>).



A&D Company, Limited

3-23-14 Higashi-Ikebukuro, Toshima-ku, Tokyo 170-0013, JAPAN
Telephone: [81] (3) 5391-6132 Fax: [81] (3) 5391-6148

A&D ENGINEERING, INC.

1756 Automation Parkway, San Jose, California 95131, U.S.A.
Telephone: [1] (408) 263-5333 Fax: [1] (408) 263-0119

A&D INSTRUMENTS LIMITED

Unit 24/26 Blacklands Way, Abingdon Business Park, Abingdon, Oxfordshire OX14 1DY United Kingdom
Telephone: [44] (1235) 550420 Fax: [44] (1235) 550485

A&D AUSTRALASIA PTY LTD

32 Dew Street, Thebarton, South Australia 5031, AUSTRALIA
Telephone: [61] (8) 8301-8100 Fax: [61] (8) 8352-7409

A&D KOREA Limited

한국에이.엔.디(주)
대한민국 서울시 영등포구 여의도동 36-2 맨하탄 빌딩 8층 우편 번호 150-749
(Manhattan Building 8th Floor, 36-2 Yoido-dong, Youngdeungpo-gu, Seoul, 150-749 Korea)
전화: [82] (2) 780-4101 팩스: [82] (2) 782-4280

OOO A&D RUS

ООО "ЭЙ энд ДИ РУС"
121357, Российская Федерация, г.Москва, ул. Верейская, дом 17
(Business-Center "Vereyskaya Plaza-2" 121357, Russian Federation, Moscow, Vereyskaya Street 17)
тел.: [7] (495) 937-33-44 факс: [7] (495) 937-55-66

A&D INSTRUMENTS INDIA PRIVATE LIMITED

ऐ&डी इन्स्ट्रुमेन्ट्स इण्डिया प्रा० लिमिटेड
509, उद्योग विहार, फेस -5, गुडगांव - 122016, हरियाणा, भारत
(509, Udyog Vihar, Phase-V, Gurgaon - 122 016, Haryana, India)
फोन : 91-124-4715555 फैक्स : 91-124-4715599